



Cizik
School of Nursing

Simulation Lab

Medical Equipment

100% Construction Documents

CD No. 045012.0001

July 02, 2018

CANNONDESIGN

Medical & Lab Equipment Planning
Houston, TX

TABLE OF CONTENTS

SECTION 1 – EXECUTIVE SUMMARY

Intent / Assumptions / Alternates / Cost Summary.....	1
Narrative / Listing of Owner Directed Changes, This Issue	2
Preliminary Budgetary Estimate, All FF&E, Sorted by Department.....	7

SECTION 2 – REPORTS CONTAINED HEREIN

Room by Room Listing of Medical Equipment Items.....	6
Summary of Medical Equipment Items, Sorted by ID.....	49
Summary of Medical Equipment Items, Sorted by Manufacturer.....	60
Product Data Cover Sheets, Sorted by Description.....	115

INTENT OF THIS REPORT

This report was created as a tool for reviewing medical equipment as it relates to this project and is not intended for use as a procurement document to indicate the actual make or model that will be selected prior to opening the new space. The accuracy of the information provided will be confirmed by the Owner and manufacturer during the equipment selection and procurement process. Final equipment selections may vary.

This issue is intended to provide the project team and owner with a CD-level workbook for coordination and equipment selection discussions. The information contained within this report is for information only, to be used as the basis of design.

Equipment pricing estimates are available for the Owner by request.

The information contained herein is considered confidential and is not available for use except in reference to the above stated project.

ASSUMPTIONS & QUALIFICATIONS NARRATIVE

CannonDesign Equipment, the Medical Equipment Consultant, was provided a pre-existing equipment budget and specifications produced by Jane Crofut Kleinmen, to utilize as the basis of design. Equipment validation meetings narrowed down quantities of existing equipment to be relocated and new equipment to be procured. Alternate manufacturers and models were discussed, as necessary.

The following is a list of departments included in this project, which references Exhibit H FFE Plan All New dated October 10, 2017 by Jane Crofut Kleinman and the CannonDesign-produced Simulation Lab Re-validation Report dated November 16, 2017.

- A. Fourth Floor Simulation Lab
 - o Administration
 - o Patient Care Flex
 - o Skills & Tasks
 - o Standardized Patients
 - o Storage

The following is a list of equipment intended for reuse in the project, per Owner feedback.

- A. Existing Equipment for reuse
- o Fetal Monitor & Cart (OB Patient Flex)
 - o Infant Scale and Instrument Table (OB Patient Flex)
 - o Resuscitation Cart (Skills & Tasks Room)
 - o ALL Exam Tables
 - o ALL Vascular Dopplers
 - o ALL Commode Chairs
 - o ALL Massage Tables
 - o ALL Transport Stretchers
 - o ALL Pulse Oximeters
 - o ALL Blood Glucose Monitors
 - o ALL Patient Transfer Boards

Five storage rooms along with associated equipment to be procured, were included in the initial equipment list referenced in Exhibit H FFE Plan All New, dated October 10, 2017 by Jane Crofut Kleinman. Two of the five storage rooms have since been removed from the scope or work. Equipment previously designated for these rooms have been relocated to remaining storage rooms, where possible but may negatively impact affect circulation space within.

All Information Systems equipment such as printers, telephones, computers, copiers, fax machines, shredders, scanners and non-medical related monitors are *excluded* from the medical equipment scope of work. They are noted and shown in plan for design and MEP coordination and only.

All lease items such as coffee service, hand sanitizer, paper towel dispensers, soap dispensers, glove box dispensers, and sharps collectors are *excluded* from the medical equipment scope of work. They are noted and shown in plan for design and MEP coordination only.

Analog clocks and standard open-top waste are noted in the equipment list for tracking purposes but are not shown in plan.

The Owner is considering an inventory management system for future inventory tracking. A wall-mounted computer workstation is shown in plan in the SIM Equipment and Bulk Storage rooms and a wall-mounted computer workstation and printer allowance are noted in the equipment list. Both are shown for future design and coordination but are *excluded* from the medical equipment scope of work.

The Home Health Apartment and SmartMed Lab (Level 5) are *excluded* from the medical equipment scope of work.

The Washer/Dryer were removed from the medical equipment scope of work, per Owner feedback.

ALTERNATE EQUIPMENT

Alternates are a means for the Owner to defer certain decisions on the scope of a project and provides an option for the Owner in determining the final scope of the project.

CannonDesign Equipment, the Medical Equipment Consultant, was asked by the Owner to provide an alternate headwall column design and alternate medication dispensing cart for their review and consideration.

- A. Alternate Item No. 1: Replace Headwall Column #4860-014

Replace Headwall Column size and change layout location

- o Alternate shown in plans – column size will change to 12" x 12" and will relocate to the patient right, offset 2'-0" from the face of the floor heaters to face of column headwall

- B. Alternate Item No. 2: Replace Medication Dispensing Cart #7099-007

Replace Medication Dispensing Cart

- o Medication dispensing cart manufacturer/model will change: NexusADC - Capsa Healthcare

COST SUMMARY

Pertaining to Documents Issued Herewith

- A. None (Equipment pricing estimates are available for the Owner by request).

Room By Room Listing of Medical Equipment

CANNONDESIGN

Medical & Lab Equipment Planning
Houston, TX

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Administration

Building: Unassigned

Room: Office, Director - Administration Room#: 484 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5177-020	1 O/O	1 Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
4988-000	0 O/O	6 Computer, Desktop		Project Draft (New)	Unassigned Unassigned	
5866-000	0 O/O	6 Printer, Laser		Project Draft (New)	Unassigned Unassigned	
6097-000	0 O/V	6 Telephone, Desktop		Project Draft (New)	Unassigned Unassigned	
4688-028	1 O/O	3 Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Existing (Reuse) Draft (New)	Unassigned Unassigned	Existing FTE's

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Administration

Building: Unassigned

Room: Office, General - Administration Room#: 482 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: 3 FTE's / 3 Desks

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5177-020	1 O/O	1 Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
4988-000	0 O/O	6 Computer, Desktop		Project Draft (New)	Unassigned Unassigned	
5866-000	0 O/O	6 Printer, Laser		Project Draft (New)	Unassigned Unassigned	Existing
6097-000	0 O/V	6 Telephone, Desktop		Project Draft (New)	Unassigned Unassigned	
4688-028	2 O/O	3 Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
4688-028	1 O/O	3 Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	New FTE

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Administration

Building: Unassigned

Room: Office, Tech Room#: 480 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: 4 FTE's / 1 Workarea

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5177-020	1 O/O	1 Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
4988-000	0 O/O	6 Computer, Desktop		Project Draft (New)	Unassigned Unassigned	
5866-000	0 O/O	6 Printer, Laser		Project Draft (New)	Unassigned Unassigned	Existing
6097-000	0 O/V	6 Telephone, Desktop		Project Draft (New)	Unassigned Unassigned	
4688-028	2 O/O	3 Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

**FKP CannonDesign
 UTHSC SON Simulation Center
 Room By Room Detail Report**

Department: Patient Care Flex

Building: Unassigned

Room: Alcove, Code Cart Room#: 227 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: Only two Crash Carts, per Eric (via Steve). Third cart has not been identified ; Crash Cart utilized as place holder for budget purposes.

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5859-085	2 O/O	3 Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Project Draft (New)	Unassigned Unassigned	
5860-013	1 O/O	3 Cart, Procedure, Resuscitation, Pediatric Lifeline LECPEDS2	InterMetro Industries Corp (LECPEDS2)	Project Draft (New)	Unassigned Unassigned	
3678-047	3 O/O	2 Defibrillator, Monitor, w/Pacing R Series ALS w/Exp Pkg, Masimo SpO2, ETCO2, NIBP	Zoll Medical Corporation (3 0720 0052 0131 0012) Zoll Medical Corporation (3 0720 0052 0131 0012)	Project Draft (New)	Unassigned Unassigned	on crash cart
3374-007	3 O/O	2 Pump, Suction/Aspirator, General, Portable SSCOR DUET (AE-6975)	Armstrong Medical Industries (AE-6975) SSCOR, Inc. (AE-6975)	Project Draft (New)	Unassigned Unassigned	on crash cart

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Control Room Room#: 470K Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: Refer to A/V plans for A/V equipment coordination This room has Sim Clock Controls for PF Rooms, per Eric 3/16/18

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6978-017	0 O/C 1	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	exisitng wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
6531-000	0 O/V 6	Telephone, Desktop, Digital, Single Line		Project Draft (New)	Unassigned Unassigned	5 desk phones (1 phone per station)
4688-028	4 O/O 0	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Debrief Room Room#: 470D, 470,G, 470H, 475 Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6978-017	0 O/C 1	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	
4988-000	0 O/O 6	Computer, Desktop		Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	exisitng wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
8149-000	0 O/V 6	Telephone, Wall, Digital, Single Line		Project Draft (New)	Unassigned Unassigned	
4688-028	1 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Nursing Alcove Room#: 470E Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: This room has Sim Clock Controls for PF Rooms, per Eric 3/16/18

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5361-029	2 O/O 2	Analyzer, Lab, Blood Gas, Point-of-Care I-STAT 1 w/Distribution Kit (Waived)	Abbott Point of Care Inc (04J60-20)	Project Draft (New)	Unassigned Unassigned	Two hospital-level POC testing analyzers requested per Kennessa/Vuong
7099-007	2 O/O 2	Cart, Medication, w/Computer Savvy Mobile Cart	Omnicell, Inc. (MMM-FRM-001)	Project Draft (New)	Unassigned Unassigned	2 designated for the entire PF area This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK
6978-017	0 O/C 1	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	
4988-000	0 O/O 6	Computer, Desktop		Project Draft (New)	Unassigned Unassigned	
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	Existing paper towel dispenser
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	Existing dispenser
3836-114	1 O/O 3	Hamper, Linen 2010A Chrome (Assembled)	Blickman Inc. (0962010100)	Project Draft (New)	Unassigned Unassigned	1 designated for the entire PF Room

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Nursing Alcove Room#: 470E Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments: This room has Sim Clock Controls for PF Rooms, per Eric 3/16/18

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6097-000	0 O/V 6	Telephone, Desktop _____	_____	Project Draft (New)	Unassigned Unassigned	
8149-000	0 O/V 6	Telephone, Wall, Digital, Single Line _____	_____	Project Draft (New)	Unassigned Unassigned	
4687-069	1 O/O 3	Waste Can, Bio-Hazardous Tough Guy 6GAJ6 (8 gal. Red)	Grainger (6GAJ6) _____	Project Draft (New)	Unassigned Unassigned	1 designated for the entire PF Room
4688-028	1 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex Corridor Room#: 470 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4071-073	4 O/O 2	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	
6529-000	0 O/V 6	Telephone, Wall		Project Draft (New)	Unassigned Unassigned	Wall phone in PF area at each entry

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, ICU Room#: 470B, 470C, 470I, 470J Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3417-098	1 O/O 2	Bed, Electric Advanta 2 [AD-100]	Hill-Rom - Bed & Stretcher Group (AD100)	Project Draft (New)	Unassigned Unassigned	
7788-038	1 O/C 1	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*) VividBoard - a GMi Company (VBINHMY23*)	Project Draft (New)	Unassigned Unassigned	
3449-000	2 O/C 1	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	two computers per bed - mannequin vitals/internal
3455-065	1 O/O 5	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	
5804-116	1 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	
BRJ872V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-146	1 O/O 3	Cart, Supply, Single Dr Room 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC)	Project Draft (New)	Unassigned Unassigned	
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount SCA Tissue North America (5510282) Tork Elevation Matic (5510282)		Lease Draft (New)	Unassigned Unassigned	Existing paper towel dispenser

FKP CannonDesign

UTHSC SON Simulation Center

Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, ICU Room#: 470B, 470C, 470I, 470J Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	Existing dispenser
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	Existing sharps 2 gallon
3806-000	1 O/O 3	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	on headwall
3803-000	1 O/O 3	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	on headwall
4860-014	2 O/C 1	Headwall Column, Floor to Ceiling Proxima	Hospital Systems, Inc. (1904239) Hospital Systems, Inc. (1904239)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accessories
3843-020	2 O/C 1	Headwall, Flatwall, 1 Patient Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159) Hospital Systems, Inc. (1904159)	Project Draft (New)	Unassigned Unassigned	
6852-001	1 C/C 1	Nurse Call, Interface, Bed TBD	Hill-Rom - Communications (TBD)	Project Draft (New)	Unassigned Unassigned	on headwall
4091-039	1 O/C 1	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77510/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77510/11710/25020)	Project Draft (New)	Unassigned Unassigned	Rail mounted for headwall
6643-003	1 O/O 2	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, ICU Room#: 470B, 470C, 470I, 470J Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4181-008	1 O/O	2 Pump, Infusion, PCA Alaris PCA Pump Module (8120)	BD - Alaris Infusion (8120)	Project Draft (New)	Unassigned Unassigned	NM
6891-005	1 O/O	2 Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
4248-000	1 O/O	3 Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	on headwall
4360-042	2 O/O	3 Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	1 IV stands designated for the Alaris Pump Controller
5934-014	1 O/O	3 Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100)	Project Draft (New)	Unassigned Unassigned	
4688-028	1 O/O	3 Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, Maternity / OB Room#: 470A Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3420-015	1 O/O 2	Bed, Electric, Birthing Affinity 4 [AF400]	Hill-Rom - Bed & Stretcher Group (P3700 / AF400)	Project Draft (New)	Unassigned Unassigned	
7788-038	1 O/C 1	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*)	Project Draft (New)	Unassigned Unassigned	
3449-000	2 O/C 1	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	two computers per bed - mannequin vitals/internal
3455-065	1 O/O 5	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	
5804-116	1 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	
5802-030	1 O/O 3	Cart, Equipment, Fetal Monitor CO-0025-62 (For Corometrics 250cx)	GCX Corporation (CO-0025-62)	Project Draft (New)	Unassigned Unassigned	cart for fetal monitor - use this to designate fetal monitor placement
BRJ873V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-146	1 O/O 3	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC)	Project Draft (New)	Unassigned Unassigned	
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, Maternity / OB Room#: 470A Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	1 designated for the entire PF Room Existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	Existing paper towel dispenser
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	Existing dispenser
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	Existing sharps 2 gallon
3806-000	1 O/O 3	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	on headwall
3803-000	1 O/O 3	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	on headwall
3843-020	1 O/C 1	Headwall, Flatwall, 1 Patient Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159) Hospital Systems, Inc. (1904159)	Project Draft (New)	Unassigned Unassigned	
5254-017	1 O/O 2	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Existing (Reuse) Draft (New)	Unassigned Unassigned	Existing - sits on cart NM (use cart for footprint)
6852-001	1 C/C 1	Nurse Call, Interface, Bed TBD	Hill-Rom - Communications (TBD)	Project Draft (New)	Unassigned Unassigned	on headwall

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, Maternity / OB Room#: 470A Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4091-039	1 O/C 1	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77510/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77510/11710/25020)	Project Draft (New)	Unassigned Unassigned	Rail mounted for headwall
6643-003	1 O/O 2	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18
4181-008	1 O/O 2	Pump, Infusion, PCA Alaris PCA Pump Module (8120)	BD - Alaris Infusion (8120)	Project Draft (New)	Unassigned Unassigned	NM
6891-005	1 O/O 2	Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
4182-008	1 O/O 2	Pump, Infusion, Syringe Alaris Syringe Module (8110)	BD - Alaris Infusion (8110)	Project Draft (New)	Unassigned Unassigned	NM
4248-000	1 O/O 3	Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	on headwall
5427-035	1 O/O 2	Scale, Clinical, Infant/Pediatric, Digital seca 334 Mobile Digital	seca Corp (334 1321 008)	Existing (Reuse) Draft (New)	Unassigned Unassigned	Sits on instrument table
4360-042	2 O/O 3	Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	1 IV stands designated for the Alaris Pump Controller
4433-003	1 O/O 3	Stretcher, Crib, Pediatric 500 Pediatric Crib/Stretcher	Pedigo Products, Inc (500-1)	Project Draft (New)	Unassigned Unassigned	specified by Jane Kleinman

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, Maternity / OB Room#: 470A Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5912-051	1 O/O 3	Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	Infant scale sits on top
5934-014	1 O/O 3	Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100) _____	Project Draft (New)	Unassigned Unassigned	
4688-028	1 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, NICU Room#: 470L Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
7788-038	1 O/C 1	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*)	Project Draft (New)	Unassigned Unassigned	
3449-000	2 O/C 1	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	two computers per bed - mannequin vitals/internal
3455-065	1 O/O 5	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	
5804-116	1 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	
5802-030	1 O/O 3	Cart, Equipment, Fetal Monitor CO-0025-62 (For Corometrics 250cx)	GCX Corporation (CO-0025-62)	Project Draft (New)	Unassigned Unassigned	cart for fetal monitor - use this to designate fetal monitor placement
BRJ874V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-146	1 O/O 3	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC)	Project Draft (New)	Unassigned Unassigned	
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount SCA Tissue North America (5510282) Tork Elevation Matic (5510282)		Lease Draft (New)	Unassigned Unassigned	Existing paper towel dispenser

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, NICU Room#: 470L Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	Existing dispenser
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	Existing sharps 2 gallon
3806-000	1 O/O 3	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	on headwall
3803-000	1 O/O 3	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	on headwall
4860-014	1 O/C 1	Headwall Column, Floor to Ceiling Proxima	Hospital Systems, Inc. (1904239) Hospital Systems, Inc. (1904239)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accessories
7249-003	1 O/O 2	Light, Treatment, Phototherapy, Infant, In-bed neoBLUE Blanket LED Phototherapy System	Natus Newborn Care (006244)	Project Draft (New)	Unassigned Unassigned	
5254-017	1 O/O 2	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Project Draft (New)	Unassigned Unassigned	Sits on cart NM (use cart for footprint)
6852-001	1 C/C 1	Nurse Call, Interface, Bed TBD	Hill-Rom - Communications (TBD)	Project Draft (New)	Unassigned Unassigned	on headwall

FKP CannonDesign

UTHSC SON Simulation Center

Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, NICU Room#: 470L Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4091-039	1 O/C	1 1 Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77510/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77510/11710/25020)	Project Draft (New)	Unassigned Unassigned	Rail mounted for headwall
6643-003	1 O/O	2 2 Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18
4181-008	1 O/O	2 2 Pump, Infusion, PCA Alaris PCA Pump Module (8120)	BD - Alaris Infusion (8120)	Project Draft (New)	Unassigned Unassigned	NM
6891-005	1 O/O	2 2 Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
4182-008	1 O/O	2 2 Pump, Infusion, Syringe Alaris Syringe Module (8110)	BD - Alaris Infusion (8110)	Project Draft (New)	Unassigned Unassigned	NM
4248-000	1 O/O	3 3 Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	on headwall
4360-042	2 O/O	3 3 Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	1 IV stands designated for the Alaris Pump Controller
5912-051	1 O/O	3 3 Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357)	Project Draft (New)	Unassigned Unassigned	
5934-014	1 O/O	3 3 Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Patient Care Flex

Building: Unassigned

Room: Patient Flex, NICU Room#: 470L Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4661-010	1 O/O 2	Warmer, Infant, Care System Giraffe OmniBed (E&W Port/N&S Grommet)	GE Healthcare - Maternal/Infant Care (6650-0004-901)	Project Draft (New)	Unassigned Unassigned	
4688-028	1 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign

UTHSC SON Simulation Center

Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (10) Room#: 440,450, 4H07 Room Sign: Area/Phase: Unassigned Room Qty: 3

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3417-099	10 O/O 2	Bed, Electric Advanta 2 [AD-130]	Hill-Rom - Bed & Stretcher Group (AD130)	Project Draft (New)	Unassigned Unassigned	Client requested ICU in thi area but costs beyond budget. Jane budgeted for MEd/Surg Advanta 2 (30 total)
5873-042	4 O/C 1	Board, White, Dry Erase Tactics Communication Board 72X36	Peter Pepper Products, Inc. (TC7236)	Project Draft (New)	Unassigned Unassigned	Placeholder for supply storage cabinet-mtd. dry erase surfaceThis is only a PLACEHOLDER to budget for the cost to customize the Solaire supply cart #5847-143 with a white board surface
3449-000	20 O/C 1	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	two computers per bed - mannequin vitals/internal
3455-065	10 O/O 5	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	
3581-045	1 O/O 3	Cart / Trolley, Mannequin, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P)	Project Draft (New)	Unassigned Unassigned	Linen Cart/Trolley & Trolley Cover for Sim-Mannequin storage
5804-116	2 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	2 per S&T Room Users requested Vendor product showcase per 4/19/18 mtg.
7099-007	1 O/O 2	Cart, Medication, w/Computer Savvy Mobile Cart	Omnicell, Inc. (MMM-FRM-001)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK
5853-066	1 O/O 3	Cart, Procedure, Isolation 7300-27Y-01-5 Isolation Cart	InnerSpace Corporation (7300-27Y-01-5)	Project Draft (New)	Unassigned Unassigned	
5859-085	1 O/O 3	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (10) Room#: 440,450, 4H07 Room Sign: Area/Phase: Unassigned Room Qty: 3

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
CS550SM	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-143	4 O/O 3	Cart, Supply, Dbl Dr Room 1.0 Glass Door Double 3210GVC	Solaire Medical (3210GVC) Solaire Medical (3210GVC)	Project Draft (New)	Unassigned Unassigned	This cart to be customized with a white board surface on the back
5025-002	1 O/O 3	Cover, Cart, Allowance TBD	InterMetro Industries Corp (TBD)	Project Draft (New)	Unassigned Unassigned	Trolley Cover for Sim-Mannequin storage
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	2 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	1 @ sink, 1 @door
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	
6145-047	1 O/O 2	Doppler, Vascular Elite 200 (2 & 5 MHz probes)	Natus Neurology - Nicolet Dopplers (ED25)	Existing (Reuse) Draft (New)	Unassigned Unassigned	NM
3806-000	10 O/O 3	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	on headwall

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (10) Room#: 440,450, 4H07 Room Sign: Area/Phase: Unassigned Room Qty: 3

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3803-000	10 O/O 3	Flowmeter, Oxygen _____	_____	Project Draft (New)	Unassigned Unassigned	on headwall
3836-114	1 O/O 3	Hamper, Linen 2010A Chrome (Assembled)	Blickman Inc. (0962010100) _____	Project Draft (New)	Unassigned Unassigned	
4860-014	5 O/C 1	Headwall Column, Floor to Ceiling Proxima	Hospital Systems, Inc. (1904239) Hospital Systems, Inc. (1904239)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accesories
3843-020	5 O/C 1	Headwall, Flatwall, 1 Patient Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159) Hospital Systems, Inc. (1904159)	Project Draft (New)	Unassigned Unassigned	
6403-027	5 O/O 2	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	NM
4071-073	5 O/O 2	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60) _____	Project Draft (New)	Unassigned Unassigned	
4091-040	10 O/C 1	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77710/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77710/11710/25020) _____	Project Draft (New)	Unassigned Unassigned	Rail mounted for both power column and headwall
4109-011	1 O/O 3	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000-159) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	NM

FKP CannonDesign

UTHSC SON Simulation Center

Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (10) Room#: 440,450, 4H07 Room Sign: Area/Phase: Unassigned Room Qty: 3

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6643-003	2 O/O 2	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	Controller plus one - 2 Modules: (2) Pump Set and one - 1 Module: (1) Pump Module- per Young 4/19/18
6891-005	1 O/O 2	Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
6891-005	1 O/O 2	Pump, Infusion, Pump Module, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
4248-000	10 O/O 3	Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	on headwall
4360-042	12 O/O 3	Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	2 IV stands designated for the Alaris Pump Controller
4414-052	2 O/O 3	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	
5934-014	10 O/O 3	Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100)	Project Draft (New)	Unassigned Unassigned	
6529-000	0 O/V 6	Telephone, Wall		Project Draft (New)	Unassigned Unassigned	@ nurse station
4688-028	12 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (5) Room#: 460 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3417-099	5 O/O 2	Bed, Electric Advanta 2 [AD-130]	Hill-Rom - Bed & Stretcher Group (AD130)	Project Draft (New)	Unassigned Unassigned	Client requested ICU in thi area but costs beyond budget. Jane budgeted for Med/Surg Advanta 2 (30 total)
5873-042	3 O/C 1	Board, White, Dry Erase Tactics Communication Board 72X36	Peter Pepper Products, Inc. (TC7236)	Project Draft (New)	Unassigned Unassigned	Placeholder for storage cabinet-mtd. dry erase surface This is only a PLACEHOLDER to budget for the cost to customize the Solaire supply cart #5847-143 with a white board surface
3449-000	10 O/C 1	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	two computers per bed - mannequin vitals/internal
3455-065	5 O/O 5	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	
3581-045	1 O/O 3	Cart / Truck, Linen, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P)	Project Draft (New)	Unassigned Unassigned	Linen Cart/Trolley & Trolley Cover for Sim-Mannequin storage
7099-007	1 O/O 2	Cart, Medication, w/Computer Savvy Mobile Cart	Omnicell, Inc. (MMM-FRM-001)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK
5853-066	1 O/O 3	Cart, Procedure, Isolation 7300-27Y-01-5 Isolation Cart	InnerSpace Corporation (7300-27Y-01-5)	Project Draft (New)	Unassigned Unassigned	
5859-085	1 O/O 3	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
BRJ870V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (5) Room#: 460 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5857-143	3 O/O 3	Cart, Supply, Dbl Dr Roam 1.0 Glass Door Double 3210GVC	Solaire Medical (3210GVC) Solaire Medical (3210GVC)	Project Draft (New)	Unassigned Unassigned	This cart to be customized with a white board surface on the back
5025-002	1 O/O 3	Cover, Cart, Allowance TBD	InterMetro Industries Corp (TBD)	Project Draft (New)	Unassigned Unassigned	Cart Cover for Linen Cart that hold the SIM MannequinsLinen Cart/Trolley & Trolley Cover for Sim-Mannequin storage
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	2 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	1 @ sink, 1 @ door
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	
6145-047	1 O/O 2	Doppler, Vascular Elite 200 (2 & 5 MHz probes)	Natus Neurology - Nicolet Dopplers (ED25)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
3806-000	5 O/O 3	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (5) Room#: 460 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3803-000	5 O/O 3	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	
3836-114	1 O/O 3	Hamper, Linen 2010A Chrome (Assembled)	Blickman Inc. (0962010100)	Project Draft (New)	Unassigned Unassigned	
4860-014	3 O/C 1	Headwall Column, Floor to Ceiling Proxima	Hospital Systems, Inc. (1904239) Hospital Systems, Inc. (1904239)	Project Draft (New)	Unassigned Unassigned	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accesories
3843-020	2 O/C 1	Headwall, Flatwall, 1 Patient Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159) Hospital Systems, Inc. (1904159)	Project Draft (New)	Unassigned Unassigned	
6403-027	5 O/O 2	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
4071-073	3 O/O 2	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	
4091-040	5 O/C 1	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77710/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77710/11710/25020)	Project Draft (New)	Unassigned Unassigned	Rail mounted for both power column and headwall
4109-011	1 O/O 3	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000-159)	Existing (Reuse) Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Skills and Tasks

Building: Unassigned

Room: Skills & Tasks, Bed (5) Room#: 460 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6643-003	2 O/O 2	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	Controller plus one - 2 Modules: (2) Pump Set and one - 1 Module: (1) Pump Module- per Young 4/19/18
6891-005	1 O/O 2	Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
6891-005	1 O/O 2	Pump, Infusion, Pump Module, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance placeholder - Pump Module not in AttaniaNM
4248-000	5 O/O 3	Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	
4360-042	7 O/O 3	Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	2 IV stands designated for the Alaris Pump Controller
4414-052	1 O/O 3	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	
5934-014	5 O/O 3	Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100)	Project Draft (New)	Unassigned Unassigned	
6529-000	0 O/V 6	Telephone, Wall		Project Draft (New)	Unassigned Unassigned	@ nurse station
4688-028	5 O/O 3	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: Exam Room Room#: 401 - 421 (Odd numbers) Room Sign: Area/Phase: Unassigned Room Qty: 10

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5804-116	1 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	
8300-009	1 O/O 0	Cart, Procedure, Accesories, Allowance TBD	InnerSpace Corporation (TBD)	Project Draft (New)	Unassigned Unassigned	Point-of-Use storage above-surface extension (to use in absence of a worksurface)
5863-324	1 O/O 3	Cart, Procedure, General 7300-27W-01-6 White w/Key Lock	InnerSpace Corporation (7300-27W-01-6)	Project Draft (New)	Unassigned Unassigned	
BRJ875V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-146	1 O/O 3	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC)	Project Draft (New)	Unassigned Unassigned	
6978-017	0 O/C 1	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	
8794-002	1 O/C 1	Diagnostic System, Integrated Green Series 777 [77791-2MPX]	Welch Allyn, Inc., A Hill-Rom Company (77791-2MPX)	Project Draft (New)	Unassigned Unassigned	ENT wall system for headwall
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: Exam Room Room#: 401 - 421 (Odd numbers) Room Sign: Area/Phase: Unassigned Room Qty: 10

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	
4109-011	1 O/O 3	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000-159)	Existing (Reuse) Draft (New)	Unassigned Unassigned	NM
4414-052	1 O/O 3	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	
9600-019	10 O/O 2	Table, Exam/Treatment, Manual Adjust, Electric Ritter 204 (Dwr Htr, Pelvic Tilt, Seamless Top)	Midmark Corporation (204-002/ 002- 0871) Midmark Corporation (204-002/ 002- 0871)	Existing (Reuse) Draft (New)	Unassigned Unassigned	Qty: 10 Existing
4920-117	1 O/O 3	Waste Can, Step-On 1883609 Slim Jim Resin Front Step 8 Gal./Black	Rubbermaid Commercial Products (1883609)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: Psy Consultation Room#: 423, 425, 427, 429 Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
5804-116	1 O/O 2	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	
8300-009	1 O/O 0	Cart, Procedure, Accesories, Allowance TBD	InnerSpace Corporation (TBD)	Project Draft (New)	Unassigned Unassigned	Point-of-Use storage above-surface extension (to use in absence of a worksurface)
5863-324	1 O/O 3	Cart, Procedure, General 7300-27W-01-6 White w/Key Lock	InnerSpace Corporation (7300-27W-01-6)	Project Draft (New)	Unassigned Unassigned	
BRJ876V	1 O/O 3	Cart, Supply, Accesories, Allowance	Solaire Medical (TBD)	Project Draft (New)	Unassigned Unassigned	Allowance for Dbl Dr Supply Cart accesories
5857-146	1 O/O 3	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC)	Project Draft (New)	Unassigned Unassigned	
6978-017	0 O/C 1	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	
8794-002	1 O/C 1	Diagnostic System, Integrated Green Series 777 [77791-2MPX]	Welch Allyn, Inc., A Hill-Rom Company (77791-2MPX)	Project Draft (New)	Unassigned Unassigned	ENT wall system for headwall
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: Psy Consultation Room#: 423, 425, 427, 429 Room Sign: Area/Phase: Unassigned Room Qty: 4

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	
4109-011	1 O/O 3	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000-159)	Existing (Reuse) Draft (New)	Unassigned Unassigned	NM
4414-052	1 O/O 3	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	
5082-017	1 O/O 2	Stretcher, Procedure, EYE/ ENT, Chair TMM5-FB Surgical Stretcher-Chair	TransMotion Medical, Inc. (TMM5-FB)	Project Draft (New)	Unassigned Unassigned	
4920-117	1 O/O 3	Waste Can, Step-On 1883609 Slim Jim Resin Front Step 8 Gal./Black	Rubbermaid Commercial Products (1883609)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: SP Corridor - Exam Room#: 4H01 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4071-073	1 O/O 2	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	
4266-051	1 O/O 2	Scale, Clinical, Adult, Digital, Floor seca 769 Digital Column Scale w/BMI	seca Corp (769 1321 994)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Standardized Patients

Building: Unassigned

Room: SP Corridor - Psy Room#: 4H02 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4071-073	1 O/O 2	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	
4266-051	1 O/O 2	Scale, Clinical, Adult, Digital, Floor seca 769 Digital Column Scale w/BMI	seca Corp (769 1321 994)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: Bulk Storage, HD Shelving Room#: 435 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
7283-002	1 O/O 3	Bin, Plastic, Allowance TBD	Akro-Mils (TBD)	Project Draft (New)	Unassigned Unassigned	
3449-000	0 O/C 1	Bracket, Computer Workstation, Wall		Project Draft (New)	Unassigned Unassigned	wall mounted computer for future inventory management
4927-026	1 O/O 3	Bracket, Patient Transfer Device, Wall Mount Wall Storage Rack TBR-2P w/ TBS- 2372 (23x72 Board)	Techno-Aide (TBS-2372 / TBR-2P)	Project Draft (New)	Unassigned Unassigned	New bracket ONLY for 5 existing patient transfer boards
8825-000	12 O/O 4	Cane, Quad Base		Project Draft (New)	Unassigned Unassigned	NM
6339-004	2 O/O 3	Cart, Supply, Linen 332 w/cover	Lakeside Manufacturing, Inc. (332)	Project Draft (New)	Unassigned Unassigned	Moved here from Skills and Tasks. It will live here and be utilized as a linen exchanged cart as needed - per meeting 3/16/18
3603-029	3 O/O 3	Chair, Clinical, Commode, Floor Foldable Commode	Medline Industries Inc. (MDS89664F)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
5177-020	1 O/O 1	Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
8821-000	12 O/O 4	Crutches, Axillary, Adult		Project Draft (New)	Unassigned Unassigned	NM
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: Bulk Storage, HD Shelving Room#: 435 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6922-001	0 O/O 3	Printer, Allowance TBD	Hewlett-Packard () _____	Project Draft (New)	Unassigned Unassigned	printer for future inventory management
4193-001	2 O/C 1	Rack, Crutch/Cane/Walker Walker Rack 7297	Performance Health (7297) _____	Project Draft (New)	Unassigned Unassigned	
4277-011	1 O/O 2	Scale, Clinical, Infant, Mobile Scale-Tronix 4802D Standard w/ 1 gm Weigh & Cart	Welch Allyn, Inc., A Hill-Rom Company (4802D-AX-XB /48800-1) _____	Project Draft (New)	Unassigned Unassigned	
4941-005	1 O/V 3	Shelving, Allowance, Supply, High Density (Movable) QuickWire QWFLTR	InnerSpace Corporation (QWFLTR (Custom)) InnerSpace Corporation (QWFLTR (Custom))	Project Draft (New)	Unassigned Unassigned	Allowance estimate for custom equipment
4298-010	4 O/O 3	Shelving, Wire, Chrome, 48 inch Quickwire QWP5247	InnerSpace Corporation (QWP5247) InnerSpace Corporation (QWP5247)	Project Draft (New)	Unassigned Unassigned	
4429-058	2 O/O 3	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	
4589-048	2 O/O 2	Ventilator, Adult / Pediatric VELA Plus	Vyair Medical (16531-00) Vyair Medical (16531-00)	Project Draft (New)	Unassigned Unassigned	*Order with rolling stand accessory, per Vuong request
6740-000	4 O/O 3	Walker, Bariatric _____	_____	Project Draft (New)	Unassigned Unassigned	QTY: 4 - 1 per ST areaNM
5407-030	1 O/O 3	Waste Can, 20-31 Gallon 3569-88 Untouchable Square (Black, 23 gal)	Rubbermaid Commercial Products (FG356988BLA) _____	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: Bulk Storage, HD Shelving Room#: 435 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
6292-017	4 O/O 3	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF) Drive DeVilbiss Healthcare (STD22ECDFA-SF)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: S&T Workroom/Storage Room#: 437 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4881-006	2 O/O 2	Bench, Tech Workstation, 6 ft. TWS-7000-ESD-0	Teclab (TWS-7000-ESD-0)	Project Draft (New)	Unassigned Unassigned	
7283-002	1 O/O 3	Bin, Plastic, Allowance TBD	Akro-Mils (TBD)	Project Draft (New)	Unassigned Unassigned	
5836-015	5 O/O 3	Cart, Utility, Polymer 4091 Xtra (Black)	Rubbermaid Commercial Products (FG409100BLA)	Project Draft (New)	Unassigned Unassigned	
5177-020	1 O/O 1	Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
5869-011	1 O/C 1	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12) GOJO Industries (9699-12)	Project Draft (New)	Unassigned Unassigned	
6084-061	1 O/C 1	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282) SCA Tissue North America (5510282)	Project Draft (New)	Unassigned Unassigned	
5868-069	1 O/C 1	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06) GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	
3723-039	1 O/C 1	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409) BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: S&T Workroom/Storage Room#: 437 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4298-010	4 O/O 3	Shelving, Wire, Chrome, 48 inch Quickwire QWP5247	InnerSpace Corporation (QWP5247) InnerSpace Corporation (QWP5247)	Project Draft (New)	Unassigned Unassigned	
7607-006	1 O/O 2	Table, Therapy, Massage, Powered Ellora Vista Lift (30"W)	EarthLite (ELLORA-30) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	
6529-000	0 O/V 6	Telephone, Wall _____	_____ _____	Project Draft (New)	Unassigned Unassigned	
5407-030	1 O/O 3	Waste Can, 20-31 Gallon 3569-88 Untouchable Square (Black, 23 gal)	Rubbermaid Commercial Products _____ (FG356988BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: SIM Equipment Storage Room#: 467 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3417-098	1 O/O 2	Bed, Electric Advanta 2 [AD-100]	Hill-Rom - Bed & Stretcher Group (AD100)	Project Draft (New)	Unassigned Unassigned	
7283-002	1 O/O 3	Bin, Plastic, Allowance TBD	Akro-Mils (TBD)	Project Draft (New)	Unassigned Unassigned	
3449-000	0 O/C 1	Bracket, Computer Workstation, Wall		Project Draft (New)	Unassigned Unassigned	wall mounted for inventory management
4927-026	1 O/O 3	Bracket, Patient Transfer Device, Wall Mount Wall Storage Rack TBR-2P w/ TBS- 2372 (23x72 Board)	Techno-Aide (TBS-2372 / TBR-2P)	Project Draft (New)	Unassigned Unassigned	
8825-000	2 O/O 4	Cane, Quad Base		Project Draft (New)	Unassigned Unassigned	2 per PF Room NM
BQA043G	1 O/O 3	Cart, Procedure, Isolation		Project Draft (New)	Unassigned Unassigned	
5859-088	1 O/O 3	Cart, Procedure, Resuscitation Lifeline LECP13000	InterMetro Industries Corp (LECP13000)	Project Draft (New)	Unassigned Unassigned	
5863-330	7 O/O 3	Cart, Task Trainers 7300-33W-01-7 White w/Key Lock	InnerSpace Corporation (7300-33W-01-7)	Project Draft (New)	Unassigned Unassigned	
5836-015	3 O/O 3	Cart, Utility, Polymer 4091 Xtra (Black)	Rubbermaid Commercial Products (FG409100BLA)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: SIM Equipment Storage Room#: 467 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I AC	Description Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
3603-029	1 O/O 3	Chair, Clinical, Commode, Floor Foldable Commode	Medline Industries Inc. (MDS89664F)	Existing (Reuse) Draft (New)	Unassigned Unassigned	
5177-020	1 O/O 1	Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	
8821-000	2 O/O 4	Crutches, Axillary, Adult		Project Draft (New)	Unassigned Unassigned	NM 2 per PF Room
7050-025	1 O/C 1	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	
6145-047	1 O/O 2	Doppler, Vascular Elite 200 (2 & 5 MHz probes)	Natus Neurology - Nicolet Dopplers (ED25)	Project Draft (New)	Unassigned Unassigned	NM
3768-079	1 O/O 2	Electrocardiograph (ECG), Interpretive Burdick ELI 250c Interpretive w/ Stand (USB)	Mortara Instrument, Inc. (BUR250C-81X / 9911-016-50)	Project Draft (New)	Unassigned Unassigned	
6403-027	1 O/O 2	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001)	Project Draft (New)	Unassigned Unassigned	NM
6922-001	0 O/O 3	Printer, Allowance TBD	Hewlett-Packard ()	Project Draft (New)	Unassigned Unassigned	printer for future inventory management
4193-001	1 O/C 1	Rack, Crutch/Cane/Walker Walker Rack 7297	Performance Health (7297)	Project Draft (New)	Unassigned Unassigned	

FKP CannonDesign
UTHSC SON Simulation Center
Room By Room Detail Report

Department: Storage

Building: Unassigned

Room: SIM Equipment Storage Room#: 467 Room Sign: Area/Phase: Unassigned Room Qty: 1

Comments:

Atta ID	Qty F/I	Description AC Model	Manufacturer Vendor	Funding Source Item Status	Custom1 Custom2	Item Notes
4298-010	4 O/O 3	Shelving, Wire, Chrome, 48 inch Quickwire QWP5247	InnerSpace Corporation (QWP5247) InnerSpace Corporation (QWP5247)	Project Draft (New)	Unassigned Unassigned	
4416-040	4 O/O 3	Stool, Step T-40 Chrome	Clinton Industries, Inc. (T-40)	Project Draft (New)	Unassigned Unassigned	
4429-058	2 O/O 3	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000)	Project Draft (New)	Unassigned Unassigned	
4919-014	1 O/O 2	Ultrasound, Imaging, Handheld Vscan	GE Healthcare - Imaging Systems (H45571CC)	Project Draft (New)	Unassigned Unassigned	NM
4589-048	4 O/O 2	Ventilator, Adult / Pediatric VELA Plus	Vyaire Medical (16531-00)	Project Draft (New)	Unassigned Unassigned	*Order with rolling stand accessory, per Vuong request
6740-000	2 O/O 3	Walker, Bariatric		Project Draft (New)	Unassigned Unassigned	NM 2 per PF Room
4657-021	1 O/O 2	Warmer, Patient, Hypothermia Bair Hugger 775	3M Infection Prevention Division (77500)	Project Draft (New)	Unassigned Unassigned	
5407-030	1 O/O 3	Waste Can, 20-31 Gallon 3569-88 Untouchable Square (Black, 23 gal)	Rubbermaid Commercial Products (FG356988BLA)	Project Draft (New)	Unassigned Unassigned	
6292-017	2 O/O 3	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF)	Project Draft (New)	Unassigned Unassigned	2 per PF Room

**Summary by Item – Sorted
by ID No. – Listing of
Medical Equipment**

CANNONDESIGN

Medical & Lab Equipment Planning
Houston, TX

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5361-029	2 O/O	Analyzer, Lab, Blood Gas, Point-of-Care I-STAT 1 w/Distribution Kit (Waived)	Abbott Point of Care Inc (04J60-20) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3417-098	5 O/O	Bed, Electric Advanta 2 [AD-100]	Hill-Rom - Bed & Stretcher Group (AD100) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3417-099	35 O/O	Bed, Electric Advanta 2 [AD-130]	Hill-Rom - Bed & Stretcher Group (AD130) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3420-015	1 O/O	Bed, Electric, Birthing Affinity 4 [AF400]	Hill-Rom - Bed & Stretcher Group (P3700 / AF400) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4881-006	2 O/O	Bench, Tech Workstation, 6 ft. TWS-7000-ESD-0	Teclab (TWS-7000-ESD-0) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
7283-002	1 O/O	Bin, Plastic, Allowance TBD	Akro-Mils (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
7283-002	2 O/O	Bin, Plastic, Allowance TBD	Akro-Mils (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
7788-038	2 O/C	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*) _____	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
7788-038	4 O/C	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*) VividBoard - a GMi Company (VBINHMY23*)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5873-042	15 O/C	Board, White, Dry Erase Tactics Communication Board 72X36	Peter Pepper Products, Inc. (TC7236) _____	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3449-000	82 O/C	Bracket, Computer Workstation, Rail _____	_____ _____	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3449-000	0 O/C	Bracket, Computer Workstation, Wall _____	_____ _____	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4927-026	2 O/O	Bracket, Patient Transfer Device, Wall Mount Wall Storage Rack TBR-2P w/ TBS-2372 (23x72 Board)	Techno-Aide (TBS-2372 / TBR-2P) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3455-065	41 O/O	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000)	Project Draft (New)	Unassigned Unassigned	5-Furniture Unassigned
8825-000	14 O/O	Cane, Quad Base		Project Draft (New)	Unassigned Unassigned	4-Instruments Unassigned
3581-045	3 O/O	Cart / Trolley, Mannequin, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3581-045	1 O/O	Cart / Truck, Linen, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5804-116	26 O/O	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5802-030	2 O/O	Cart, Equipment, Fetal Monitor CO-0025-62 (For Corometrics 250cx)	GCX Corporation (CO-0025-62)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
7099-007	6 O/O	Cart, Medication, w/Computer Savvy Mobile Cart	Omniceil, Inc. (MMM-FRM-001)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
8300-009	14 O/O	Cart, Procedure, Accesories, Allowance TBD	InnerSpace Corporation (TBD)	Project Draft (New)	Unassigned Unassigned	0-Unassigned Unassigned
5863-324	14 O/O	Cart, Procedure, General 7300-27W-01-6 White w/Key Lock	InnerSpace Corporation (7300-27W-01-6)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5853-066	4 O/O	Cart, Procedure, Isolation 7300-27Y-01-5 Isolation Cart	InnerSpace Corporation (7300-27Y-01-5)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BQA043G	1 O/O	Cart, Procedure, Isolation		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5859-085	1 O/O	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5859-085	5 O/O	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5859-088	1 O/O	Cart, Procedure, Resuscitation Lifeline LECP13000	InterMetro Industries Corp (LECP13000) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5860-013	1 O/O	Cart, Procedure, Resuscitation, Pediatric Lifeline LECPEDS2	InterMetro Industries Corp (LECPEDS2) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ870V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ872V	4 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ873V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ874V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ875V	10 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ876V	4 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
CS550SM	3 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5857-143	15 O/O	Cart, Supply, Dbl Dr Roam 1.0 Glass Door Double 3210GVC	Solaire Medical (3210GVC) Solaire Medical (3210GVC)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6339-004	2 O/O	Cart, Supply, Linen 332 w/cover	Lakeside Manufacturing, Inc. (332) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5857-146	20 O/O	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5863-330	7 O/O	Cart, Task Trainers 7300-33W-01-7 White w/Key Lock	InnerSpace Corporation (7300-33W-01-7) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5836-015	8 O/O	Cart, Utility, Polymer 4091 Xtra (Black)	Rubbermaid Commercial Products (FG409100BLA)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3603-029	4 O/O	Chair, Clinical, Commode, Floor Foldable Commode	Medline Industries Inc. (MDS89664F)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5177-020	6 O/O	Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
6978-017	0 O/C	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4988-000	0 O/O	Computer, Desktop		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
5025-002	4 O/O	Cover, Cart, Allowance TBD	InterMetro Industries Corp (TBD)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
8821-000	14 O/O	Crutches, Axillary, Adult		Project Draft (New)	Unassigned Unassigned	4-Instruments Unassigned
3678-047	3 O/O	Defibrillator, Monitor, w/Pacing R Series ALS w/Exp Pkg, Masimo SpO2, ETCO2, NIBP	Zoll Medical Corporation (3 0720 0052 0131 0012) Zoll Medical Corporation (3 0720 0052 0131 0012)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
8794-002	14 O/C	Diagnostic System, Integrated Green Series 777 [77791-2MPX]	Welch Allyn, Inc., A Hill-Rom Company (77791- 2MPX)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
7050-025	28 O/C	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5869-011	29 O/C	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5869-011	1 O/C	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12) GOJO Industries (9699-12)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
6084-061	25 O/C	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6084-061	1 O/C	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282) SCA Tissue North America (5510282)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5868-069	25 O/C	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5868-069	1 O/C	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06) GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3723-039	24 O/C	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3723-039	1 O/C	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409) BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
6145-047	4 O/O	Doppler, Vascular Elite 200 (2 & 5 MHz probes)	Natus Neurology - Nicolet Dopplers (ED25)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6145-047	1 O/O	Doppler, Vascular Elite 200 (2 & 5 MHz probes)	Natus Neurology - Nicolet Dopplers (ED25)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3768-079	1 O/O	Electrocardiograph (ECG), Interpretive Burdick ELI 250c Interpretive w/ Stand (USB)	Mortara Instrument, Inc. (BUR250C-81X / 9911-016-50)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3806-000	41 O/O	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3803-000	41 O/O	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3836-114	5 O/O	Hamper, Linen 2010A Chrome (Assembled)	Blickman Inc. (0962010100)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4860-014	27 O/C	Headwall Column, Floor to Ceiling Proxima	Hospital Systems, Inc. (1904239) Hospital Systems, Inc. (1904239)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3843-020	26 O/C	Headwall, Flatwall, 1 Patient Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159) Hospital Systems, Inc. (1904159)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7249-003	1 O/O	Light, Treatment, Phototherapy, Infant, In-bed neoBLUE Blanket LED Phototherapy System	Natus Newborn Care (006244)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6403-027	20 O/O	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6403-027	1 O/O	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5254-017	1 O/O	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5254-017	1 O/O	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4071-073	24 O/O	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6852-001	6 C/C	Nurse Call, Interface, Bed TBD	Hill-Rom - Communications (TBD)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4091-039	6 O/C	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77510/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77510/11710/25020)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4091-040	35 O/C	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77710/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77710/11710/25020)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4109-011	18 O/O	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000- 159)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6922-001	0 O/O	Printer, Allowance TBD	Hewlett-Packard ()	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5866-000	0 O/O	Printer, Laser		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6643-003	14 O/O	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4181-008	6 O/O	Pump, Infusion, PCA Alaris PCA Pump Module (8120)	BD - Alaris Infusion (8120)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6891-005	10 O/O	Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6891-005	4 O/O	Pump, Infusion, Pump Module, Allowance TBD	BD - Alaris Infusion (TBD)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4182-008	2 O/O	Pump, Infusion, Syringe Alaris Syringe Module (8110)	BD - Alaris Infusion (8110)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
3374-007	3 O/O	Pump, Suction/Aspirator, General, Portable SSCOR DUET (AE-6975)	Armstrong Medical Industries (AE-6975) SSCOR, Inc. (AE-6975)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4193-001	3 O/C	Rack, Crutch/Cane/Walker Walker Rack 7297	Performance Health (7297)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4248-000	41 O/O	Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4266-051	2 O/O	Scale, Clinical, Adult, Digital, Floor seca 769 Digital Column Scale w/BMI	seca Corp (769 1321 994)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4277-011	1 O/O	Scale, Clinical, Infant, Mobile Scale-Tronix 4802D Standard w/ 1 gm Weigh & Cart	Welch Allyn, Inc., A Hill-Rom Company (4802D-AX- XB /48800-1)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5427-035	1 O/O	Scale, Clinical, Infant/Pediatric, Digital seca 334 Mobile Digital	seca Corp (334 1321 008)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4941-005	1 O/V	Shelving, Allowance, Supply, High Density (Movable) QuickWire QWFLTR	InnerSpace Corporation (QWFLTR (Custom)) InnerSpace Corporation (QWFLTR (Custom))	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4298-010	12 O/O	Shelving, Wire, Chrome, 48 inch Quickwire QWP5247	InnerSpace Corporation (QWP5247) InnerSpace Corporation (QWP5247)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4360-042	55 O/O	Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4414-052	21 O/O	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4416-040	4 O/O	Stool, Step T-40 Chrome	Clinton Industries, Inc. (T-40)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4433-003	1 O/O	Stretcher, Crib, Pediatric 500 Pediatric Crib/Stretcher	Pedigo Products, Inc (500-1)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5082-017	4 O/O	Stretcher, Procedure, EYE/ ENT, Chair TMM5-FB Surgical Stretcher-Chair	TransMotion Medical, Inc. (TMM5-FB)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4429-058	2 O/O	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4429-058	2 O/O	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
9600-019	100 O/O	Table, Exam/Treatment, Manual Adjust, Electric Ritter 204 (Dwr Htr, Pelvic Tilt, Seamless Top)	Midmark Corporation (204-002/ 002-0871) Midmark Corporation (204-002/ 002-0871)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5912-051	1 O/O	Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5912-051	1 O/O	Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5934-014	41 O/O	Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
7607-006	1 O/O	Table, Therapy, Massage, Powered Ellora Vista Lift (30"W)	EarthLite (ELLORA-30)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6097-000	0 O/V	Telephone, Desktop		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6531-000	0 O/V	Telephone, Desktop, Digital, Single Line _____	_____	Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
6529-000	0 O/V	Telephone, Wall _____	_____	Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
8149-000	0 O/V	Telephone, Wall, Digital, Single Line _____	_____	Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
4919-014	1 O/O	Ultrasound, Imaging, Handheld Vscan	GE Healthcare - Imaging Systems (H45571CC) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4589-048	4 O/O	Ventilator, Adult / Pediatric VELA Plus	Vyaire Medical (16531-00) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4589-048	2 O/O	Ventilator, Adult / Pediatric VELA Plus	Vyaire Medical (16531-00) Vyaire Medical (16531-00)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6740-000	6 O/O	Walker, Bariatric _____	_____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4661-010	1 O/O	Warmer, Infant, Care System Giraffe OmniBed (E&W Port/N&S Grommet)	GE Healthcare - Maternal/Infant Care (6650-0004-901) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4657-021	1 O/O	Warmer, Patient, Hypothermia Bair Hugger 775	3M Infection Prevention Division (77500) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5407-030	3 O/O	Waste Can, 20-31 Gallon 3569-88 Untouchable Square (Black, 23 gal)	Rubbermaid Commercial Products (FG356988BLA) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4687-069	1 O/O	Waste Can, Bio-Hazardous Tough Guy 6GAJ6 (8 gal. Red)	Grainger (6GAJ6) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4688-028	4 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	0-Unassigned Unassigned
4688-028	3 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4688-028	55 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4920-117	14 O/O	Waste Can, Step-On 1883609 Slim Jim Resin Front Step 8 Gal./Black	Rubbermaid Commercial Products (1883609) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6292-017	2 O/O	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6292-017	4 O/O	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF) Drive DeVilbiss Healthcare (STD22ECDFA-SF)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

**Summary by Item – Sorted
by Manufacturer. – Listing
of Medical Equipment**

CANNONDESIGN

Medical & Lab Equipment Planning
Houston, TX

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: Unspecified (Phone:)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3449-000	82 O/C	Bracket, Computer Workstation, Rail		Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3449-000	0 O/C	Bracket, Computer Workstation, Wall		Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
8825-000	14 O/O	Cane, Quad Base		Project Draft (New)	Unassigned Unassigned	4-Instruments Unassigned
BQA043G	1 O/O	Cart, Procedure, Isolation		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4988-000	0 O/O	Computer, Desktop		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
8821-000	14 O/O	Crutches, Axillary, Adult		Project Draft (New)	Unassigned Unassigned	4-Instruments Unassigned
3806-000	41 O/O	Flowmeter, Air		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3803-000	41 O/O	Flowmeter, Oxygen		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5866-000	0 O/O	Printer, Laser		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
4248-000	41 O/O	Regulator, Suction, Intermittent/Continuous		Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6097-000	0 O/V	Telephone, Desktop		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
6531-000	0 O/V	Telephone, Desktop, Digital, Single Line		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
6529-000	0 O/V	Telephone, Wall		Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: Unspecified (Phone:)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
8149-000	0 O/V	Telephone, Wall, Digital, Single Line _____	_____ _____	Project Draft (New)	Unassigned Unassigned	6-IT/Computers Unassigned
6740-000	6 O/O	Walker, Bariatric _____	_____ _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: 3M Infection Prevention Division (Phone: 952-947-1200)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4657-021	1 O/O	Warmer, Patient, Hypothermia Bair Hugger 775	3M Infection Prevention Division (77500)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Abbott Point of Care Inc (Phone: 609.443.9300)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5361-029	2 O/O	Analyzer, Lab, Blood Gas, Point-of-Care I-STAT 1 w/Distribution Kit (Waived)	Abbott Point of Care Inc (04J60-20)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Akro-Mils (Phone: (800) 253-2467)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7283-002	1 O/O	Bin, Plastic, Allowance TBD	Akro-Mils (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
7283-002	2 O/O	Bin, Plastic, Allowance TBD	Akro-Mils (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Armstrong Medical Industries (Phone: (800) 323-4220)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3374-007	3 O/O	Pump, Suction/Aspirator, General, Portable SSCOR DUET (AE-6975)	Armstrong Medical Industries (AE-6975) SSCOR, Inc. (AE-6975)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: BD - Alaris Infusion (Phone: (800) 854-7128)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6643-003	14 O/O	Pump, Infusion, Controller, Modular Alaris PC Unit (8015)	BD - Alaris Infusion (8015) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4182-008	2 O/O	Pump, Infusion, Syringe Alaris Syringe Module (8110)	BD - Alaris Infusion (8110) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4181-008	6 O/O	Pump, Infusion, PCA Alaris PCA Pump Module (8120)	BD - Alaris Infusion (8120) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6891-005	10 O/O	Pump, Infusion, Pump Module Set, Allowance TBD	BD - Alaris Infusion (TBD) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6891-005	4 O/O	Pump, Infusion, Pump Module, Allowance TBD	BD - Alaris Infusion (TBD) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: BD Medical (Phone: 800.237.2174)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3723-039	24 O/C	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409) _____	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
3723-039	1 O/C	Disposal, Sharps, Wall Mount Next Generation 305435 (2 Gal) w/ 305409 Bracket	BD Medical (305435/305409) BD Medical (305435/305409)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Blickman Inc. (Phone: 201-909-0807)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4360-042	55 O/O	Stand, IV, Stainless Steel 1415SS-4 (5-Leg, 4-Hook)	Blickman Inc. (0531415400) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3836-114	5 O/O	Hamper, Linen 2010A Chrome (Assembled)	Blickman Inc. (0962010100) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Chicago Lighthouse Industries (Phone: 800-842-8624)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5177-020	6 O/O	Clock, Analog, Wall 24 Hour Series [67300040]	Chicago Lighthouse Industries (67300040)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Clinton Industries, Inc. (Phone: 800-441-9131)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4416-040	4 O/O	Stool, Step T-40 Chrome	Clinton Industries, Inc. (T-40)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Drive DeVilbiss Healthcare (Phone: 877-224-0946)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6292-017	2 O/O	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
6292-017	4 O/O	Wheelchair, Adult, Bariatric Sentra EC 22in. [STD22ECDFA-SF]	Drive DeVilbiss Healthcare (STD22ECDFA-SF) Drive DeVilbiss Healthcare (STD22ECDFA-SF)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: EarthLite (Phone: 800-872-0560)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7607-006	1 O/O	Table, Therapy, Massage, Powered Ellora Vista Lift (30"W)	EarthLite (ELLORA-30)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: GCX Corporation (Phone: (800) 228-2555)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5802-030	2 O/O	Cart, Equipment, Fetal Monitor CO-0025-62 (For Corometrics 250cx)	GCX Corporation (CO-0025-62)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: GE Healthcare - Anesthesia Delivery (Phone: (800) 345-2700)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4109-011	18 O/O	Oximeter, Pulse, Hand Held TuffSat	GE Healthcare - Anesthesia Delivery (6051-0000-159)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: GE Healthcare - Imaging Systems (Phone: 800-886-0815)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4919-014	1 O/O	Ultrasound, Imaging, Handheld Vscan	GE Healthcare - Imaging Systems (H45571CC)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: GE Healthcare - Maternal/Infant Care (Phone: (608) 221-1551)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4661-010	1 O/O	Warmer, Infant, Care System Giraffe OmniBed (E&W Port/N&S Grommet)	GE Healthcare - Maternal/Infant Care (6650-0004-901)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: GE Healthcare - Monitoring Systems (Phone: 800-558-5120)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5254-017	1 O/O	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
5254-017	1 O/O	Monitor, O.B., Intrapartum, Maternal/Fetal Corometrics 259cx	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: GOJO Industries (Phone: 800.321.9647)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5868-069	25 O/C	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5868-069	1 O/C	Dispenser, Soap, Wall Mount GOJO FMX-12 Black (5155-06)	GOJO Industries (5155-06) GOJO Industries (5155-06)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5869-011	29 O/C	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
5869-011	1 O/C	Dispenser, Hand Sanitizer, Wall Mount Purell APX Aerosol (9699-12)	GOJO Industries (9699-12) GOJO Industries (9699-12)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Grainger (Phone: 847-535-1000)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4687-069	1 O/O	Waste Can, Bio-Hazardous Tough Guy 6GAJ6 (8 gal. Red)	Grainger (6GAJ6)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Health Care Logistics (Phone: (800) 848-1633)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7050-025	28 O/C	Dispenser, Glove, Quadruple Box 17899 Stainless Steel	Health Care Logistics (17899)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Hewlett-Packard (Phone: 650-857-1501)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6922-001	0 O/O	Printer, Allowance TBD	Hewlett-Packard () _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: Hill-Rom - Bed & Stretcher Group (Phone: 800-445-3730)

Atta ID	Qty	Description	Manufacturer (Mfr #)	Funding Source	Cost Center	Arch Code
	F/I	Model	Vendor (Vendor #)	Item Status	Budget Name	Custom Code
3417-098	5	Bed, Electric	Hill-Rom - Bed & Stretcher Group (AD100)	Project	Unassigned	2-Movable, Elect
	O/O	Advanta 2 [AD-100]	_____	Draft (New)	Unassigned	Unassigned
3417-099	35	Bed, Electric	Hill-Rom - Bed & Stretcher Group (AD130)	Project	Unassigned	2-Movable, Elect
	O/O	Advanta 2 [AD-130]	_____	Draft (New)	Unassigned	Unassigned
3420-015	1	Bed, Electric, Birthing	Hill-Rom - Bed & Stretcher Group (P3700 / AF400)	Project	Unassigned	2-Movable, Elect
	O/O	Affinity 4 [AF400]	_____	Draft (New)	Unassigned	Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Hill-Rom - Communications (Phone: 800.445.3730)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6852-001	6 C/C	Nurse Call, Interface, Bed TBD	Hill-Rom - Communications (TBD)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Hospital Systems, Inc. (Phone: 925.427.7800)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3843-020	26	Headwall, Flatwall, 1 Patient	Hospital Systems, Inc. (1904159)	Project	Unassigned	1-Fixed
	O/C	Axiom Surface Mounted Headwall	Hospital Systems, Inc. (1904159)	Draft (New)	Unassigned	Unassigned
4860-014	27	Headwall Column, Floor to Ceiling	Hospital Systems, Inc. (1904239)	Project	Unassigned	1-Fixed
	O/C	Proxima	Hospital Systems, Inc. (1904239)	Draft (New)	Unassigned	Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: InnerSpace Corporation (Phone: (800) 467-7224)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5863-324	14 O/O	Cart, Procedure, General 7300-27W-01-6 White w/Key Lock	InnerSpace Corporation (7300-27W-01-6)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5853-066	4 O/O	Cart, Procedure, Isolation 7300-27Y-01-5 Isolation Cart	InnerSpace Corporation (7300-27Y-01-5)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5863-330	7 O/O	Cart, Task Trainers 7300-33W-01-7 White w/Key Lock	InnerSpace Corporation (7300-33W-01-7)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4941-005	1 O/V	Shelving, Allowance, Supply, High Density (Movable) QuickWire QWFLTR	InnerSpace Corporation (QWFLTR (Custom)) InnerSpace Corporation (QWFLTR (Custom))	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4298-010	12 O/O	Shelving, Wire, Chrome, 48 inch Quickwire QWP5247	InnerSpace Corporation (QWP5247) InnerSpace Corporation (QWP5247)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
8300-009	14 O/O	Cart, Procedure, Accesories, Allowance TBD	InnerSpace Corporation (TBD)	Project Draft (New)	Unassigned Unassigned	0-Unassigned Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Innovation Wireless (Phone: 800-732-8950)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6978-017	0 O/C	Clock, Digital, Synchronized, Wireless Digital Display Clock 642307	Innovation Wireless (642307)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: InterMetro Industries Corp (Phone: 800 992-1776)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5859-085	1 O/O	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5859-085	5 O/O	Cart, Procedure, Resuscitation Lifeline LECCRP2	InterMetro Industries Corp (LECCRP2) InterMetro Industries Corp (LECCRP2)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5859-088	1 O/O	Cart, Procedure, Resuscitation Lifeline LECP13000	InterMetro Industries Corp (LECP13000) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5860-013	1 O/O	Cart, Procedure, Resuscitation, Pediatric Lifeline LECPEDS2	InterMetro Industries Corp (LECPEDS2) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5025-002	4 O/O	Cover, Cart, Allowance TBD	InterMetro Industries Corp (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: Lakeside Manufacturing, Inc. (Phone: (800) 558-8565)

Atta ID	Qty	Description	Manufacturer (Mfr #)	Funding Source	Cost Center	Arch Code
	F/I	Model	Vendor (Vendor #)	Item Status	Budget Name	Custom Code
6339-004	2 O/O	Cart, Supply, Linen 332 w/cover	Lakeside Manufacturing, Inc. (332) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5912-051	1 O/O	Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5912-051	1 O/O	Table, Instrument, 30-36 inch 8357 (36in.W x 24in.D)	Lakeside Manufacturing, Inc. (8357) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: MAC Medical, Inc. (Phone: 877-828-9975)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4429-058	2 O/O	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4429-058	2 O/O	Stretcher, Transport PTF1000 (5th Wheel Transport Stretcher)	MAC Medical, Inc. (PTF1000) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Medline Industries Inc. (Phone: 1-800-MEDLINE)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3603-029	4 O/O	Chair, Clinical, Commode, Floor Foldable Commode	Medline Industries Inc. (MDS89664F)	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Midmark Corporation (Phone: (800) 643-6275)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
9600-019	100 O/O	Table, Exam/Treatment, Manual Adjust, Electric Ritter 204 (Dwr Htr, Pelvic Tilt, Seamless Top)	Midmark Corporation (204-002/ 002-0871) Midmark Corporation (204-002/ 002-0871)	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4414-052	21 O/O	Stool, Exam, Cushion-Seat Ritter 272 Value Series	Midmark Corporation (272-001) Midmark Corporation (272-001)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: MODRoto Inc. (Phone: (800) 829-4535)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3581-045	3 O/O	Cart / Trolley, Mannequin, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3581-045	1 O/O	Cart / Truck, Linen, Bulk 101P (46.5 cu.ft.)	MODRoto Inc. (101P) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Mortara Instrument, Inc. (Phone: 800-231-7437)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3768-079	1 O/O	Electrocardiograph (ECG), Interpretive Burdick ELI 250c Interpretive w/ Stand (USB)	Mortara Instrument, Inc. (BUR250C-81X / 9911-016-50)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign
UTHSC SON Simulation Center

Item Summary

Manufacturer: Natus Neurology - Nicolet Dopplers (Phone: 608-441-2266)

Atta ID	Qty	Description	Manufacturer (Mfr #)	Funding Source	Cost Center	Arch Code
	F/I	Model	Vendor (Vendor #)	Item Status	Budget Name	Custom Code
6145-047	4	Doppler, Vascular	Natus Neurology - Nicolet Dopplers (ED25)	Existing (Reuse)	Unassigned	2-Movable, Elect
	O/O	Elite 200 (2 & 5 MHz probes)	_____	Draft (New)	Unassigned	Unassigned
6145-047	1	Doppler, Vascular	Natus Neurology - Nicolet Dopplers (ED25)	Project	Unassigned	2-Movable, Elect
	O/O	Elite 200 (2 & 5 MHz probes)	_____	Draft (New)	Unassigned	Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Natus Newborn Care (Phone: (800) 426-0353)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7249-003	1 O/O	Light, Treatment, Phototherapy, Infant, In-bed neoBLUE Blanket LED Phototherapy System	Natus Newborn Care (006244)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Omnicell, Inc. (Phone: (650) 251-6100)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7099-007	6 O/O	Cart, Medication, w/Computer Savvy Mobile Cart	Omnicell, Inc. (MMM-FRM-001)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Pedigo Products, Inc (Phone: (360) 695-3500)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4433-003	1 O/O	Stretcher, Crib, Pediatric 500 Pediatric Crib/Stretcher	Pedigo Products, Inc (500-1)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Performance Health (Phone: (630) 226-1300)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4193-001	3 O/C	Rack, Crutch/Cane/Walker Walker Rack 7297	Performance Health (7297)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Peter Pepper Products, Inc. (Phone: (310) 639-0390)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5873-042	15	Board, White, Dry Erase	Peter Pepper Products, Inc. (TC7236)	Project	Unassigned	1-Fixed
	O/C	Tactics Communication Board 72X36		Draft (New)	Unassigned	Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Roche Diagnostics Corporation (Phone: 800-428-5076)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6403-027	20 O/O	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
6403-027	1 O/O	Monitor, Blood Glucose, Point-of-Care ACCU-CHEK Performa Meter	Roche Diagnostics Corporation (05959551001) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Rubbermaid Commercial Products (Phone: 800-347-9800)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4920-117	14 O/O	Waste Can, Step-On 1883609 Slim Jim Resin Front Step 8 Gal./Black	Rubbermaid Commercial Products (1883609) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4688-028	3 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4688-028	55 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
4688-028	4 O/O	Waste Can, Open Top 2543 Fire Resistant Black (28 qt.)	Rubbermaid Commercial Products (FG254300BLA) _____	Project Draft (New)	Unassigned Unassigned	0-Unassigned Unassigned
5407-030	3 O/O	Waste Can, 20-31 Gallon 3569-88 Untouchable Square (Black, 23 gal)	Rubbermaid Commercial Products (FG356988BLA) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5836-015	8 O/O	Cart, Utility, Polymer 4091 Xtra (Black)	Rubbermaid Commercial Products (FG409100BLA) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Rubbermaid Healthcare, A Division of Capsa Solutions (Phone: 888-859-8294)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5804-116	26 O/O	Cart, Computer, Workstation M38e Computer Cart (40 Amp AC)	Rubbermaid Healthcare, A Division of Capsa Solutions (1781127)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: SCA Tissue North America (Phone: 866-722-8675)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
6084-061	25 O/C	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282)	Lease Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
6084-061	1 O/C	Dispenser, Paper Towel, Surface Mount Tork Elevation Matic (5510282)	SCA Tissue North America (5510282) SCA Tissue North America (5510282)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: seca Corp (Phone: 410 694 9330)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5427-035	1 O/O	Scale, Clinical, Infant/Pediatric, Digital seca 334 Mobile Digital	seca Corp (334 1321 008) _____	Existing (Reuse) Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4266-051	2 O/O	Scale, Clinical, Adult, Digital, Floor seca 769 Digital Column Scale w/BMI	seca Corp (769 1321 994) _____	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Solaire Medical (Phone:)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5857-143	15 O/O	Cart, Supply, Dbl Dr Roam 1.0 Glass Door Double 3210GVC	Solaire Medical (3210GVC) Solaire Medical (3210GVC)	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
5857-146	20 O/O	Cart, Supply, Single Dr Roam 0.25 Glass Door Single S3225RGVC	Solaire Medical (S3225RGVC) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ870V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ872V	4 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ873V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ874V	1 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ875V	10 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
BRJ876V	4 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
CS550SM	3 O/O	Cart, Supply, Accesories, Allowance _____	Solaire Medical (TBD) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Stryker Medical (Phone: 800-869-0770)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5934-014	41 O/O	Table, Overbed, General Tru-Fit Single Top w/o Vanity	Stryker Medical (3150-000-100) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned
3455-065	41 O/O	Cabinet, Patient Room, Bedside Contemporary MedSurg - 1 Dwr/1 Dr	Stryker Medical (4400-800-000) _____	Project Draft (New)	Unassigned Unassigned	5-Furniture Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Techno-Aide (Phone: (800) 251-2629)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4927-026	2 O/O	Bracket, Patient Transfer Device, Wall Mount Wall Storage Rack TBR-2P w/ TBS-2372 (23x72 Board)	Techno-Aide (TBS-2372 / TBR-2P) _____	Project Draft (New)	Unassigned Unassigned	3-Movable, Non-Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Teclab (Phone: (800) 832-5227)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4881-006	2 O/O	Bench, Tech Workstation, 6 ft. TWS-7000-ESD-0	Teclab (TWS-7000-ESD-0)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: TransMotion Medical, Inc. (Phone: 866-860-8447)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
5082-017	4 O/O	Stretcher, Procedure, EYE/ ENT, Chair TMM5-FB Surgical Stretcher-Chair	TransMotion Medical, Inc. (TMM5-FB)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: VividBoard - a GMi Company (Phone:)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
7788-038	2 O/C	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*) _____	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
7788-038	4 O/C	Board, Patient Information InHarmony Changeable Glassboard 24"x36"	VividBoard - a GMi Company (VBINHMY23*) VividBoard - a GMi Company (VBINHMY23*)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Vyair Medical (Phone: 760-778-7200)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4589-048	4 O/O	Ventilator, Adult / Pediatric VELA Plus	Vyair Medical (16531-00)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4589-048	2 O/O	Ventilator, Adult / Pediatric VELA Plus	Vyair Medical (16531-00) Vyair Medical (16531-00)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company (Phone: (800) 535-6663)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
4277-011	1 O/O	Scale, Clinical, Infant, Mobile Scale-Tronix 4802D Standard w/ 1 gm Weigh & Cart	Welch Allyn, Inc., A Hill-Rom Company (4802D-AX- XB /48800-1)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4071-073	24 O/O	Monitor, Physiologic, Vital Signs, w/Stand Connex 6800 Wireless w/ Stand	Welch Allyn, Inc., A Hill-Rom Company (68MXTX-B / 4800-60)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned
4091-039	6 O/C	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77510/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77510/11710/25020)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
4091-040	35 O/C	Oto/Ophthalmoscope Set, Rail Mount Green Series 777 [77710/11710/25020]	Welch Allyn, Inc., A Hill-Rom Company (77710/11710/25020)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned
8794-002	14 O/C	Diagnostic System, Integrated Green Series 777 [77791-2MPX]	Welch Allyn, Inc., A Hill-Rom Company (77791- 2MPX)	Project Draft (New)	Unassigned Unassigned	1-Fixed Unassigned

FKP CannonDesign

UTHSC SON Simulation Center

Item Summary

Manufacturer: Zoll Medical Corporation (Phone: 978-421-9655)

Atta ID	Qty F/I	Description Model	Manufacturer (Mfr #) Vendor (Vendor #)	Funding Source Item Status	Cost Center Budget Name	Arch Code Custom Code
3678-047	3 O/O	Defibrillator, Monitor, w/Pacing R Series ALS w/Exp Pkg, Masimo SpO2, ETCO2, NIBP	Zoll Medical Corporation (3 0720 0052 0131 0012) Zoll Medical Corporation (3 0720 0052 0131 0012)	Project Draft (New)	Unassigned Unassigned	2-Movable, Elect Unassigned

Tech Data Coversheets – Sorted by Equipment Description

CANNONDESIGN

Medical & Lab Equipment Planning
Houston, TX

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5361-029



Description: Analyzer, Lab, Blood Gas, Point-of-Care

Manufacturer: Abbott Point of Care Inc ((609) 454-9000)

Mfr #: 04J60-20

Vendor: Unspecified

Vendor #:

Model: I-STAT 1 w/Distribution Kit (Waived)

Handheld blood gas analyzer with network downloader/recharger. Utilizes self-contained test cartridges for testing Blood Gases, Electrolytes, Chemistries, Coagulation, Hematology, Glucose, and Cardiac Markers (cTnI). CLIA-waived distribution kit includes analyzer, downloader/recharger, printer, electronic simulator and manual.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 3.25 in (83 mm)	Left: N/A
Depth: 9.25 in (235 mm)	Right: N/A
Height: 3.00 in (76 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Dimensions are for analyzer only. Kit also includes Downloader/Recharger (4.12"W x 10.25"L x 5"H) and mini printer.

Unit requires remote data connection to upload patient data.

Uses two 9-volt lithium batteries or rechargeable battery.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	2	Two hospital-level POC testing analyzers requested per Kennessa/Vuong
Total:					2	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3417-098



Description: Bed, Electric

Manufacturer: Hill-Rom - Bed & Stretcher Group (800-445-3730)

Mfr #: AD100

Vendor: Unspecified

Vendor #:

Model: Advanta 2 [AD-100]

Electric Med-Surg Bed. Base model. Standard features include: SlideGuard Frame technology, Audible Brake Alarm, 15.75" Low Chair Deck height, "Dampened" Instant CPR, Smart Bed Ready (Sidecom Required), Battery back-up, Lo-Lo Indicator, Electric Vascular Foot, Night Light, 4 Dual Locking Casters and Electronic Trendelenburg. Frame Only; Surface sold separately.

General Product Detail

Arch Sig: Yes	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 636
Hz: 60	Amps: 5.3
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 42.75 in (1086 mm)	Left: N/A
Depth: 94.00 in (2388 mm)	Right: N/A
Height: 48.00 in (1219 mm)	Front: N/A
Max Weight: 465 lbs (210.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Width reflects side-rails up.
Height includes headboard.
Depth reflects bed fully extended with footboard.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	
					Total:	5

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3417-099



Description: Bed, Electric

Manufacturer: Hill-Rom - Bed & Stretcher Group (800-445-3730)

Mfr #: AD130

Vendor: Unspecified

Vendor #:

Model: Advanta 2 [AD-130]

Electric Med-Surg Bed. Base model with scale and single mode Bed Exit. Standard features include: SlideGuard Frame technology, Audible Brake Alarm, 15.75" Low Chair Deck height, "Dampened" Instant CPR, Smart Bed Ready (Sidecom Required), Battery back-up, Lo-Lo Indicator, Electric Vascular Foot, Night Light, 4 Dual Locking Casters and Electronic Trendelenburg. Frame Only; Surface sold separately.

General Product Detail

Arch Sig: Yes	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 636
Hz: 60	Amps: 5.3
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 42.75 in (1086 mm)	Left: N/A
Depth: 94.00 in (2388 mm)	Right: N/A
Height: 48.00 in (1219 mm)	Front: N/A
Max Weight: 465 lbs (210.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Width reflects side-rails up.
Height includes headboard.
Depth reflects bed fully extended with footboard.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	Client requested ICU in thi area but costs beyond budget. Jane budgeted for MEd/Surg Advanta 2 (30 total)
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	Client requested ICU in thi area but costs beyond budget. Jane budgeted for Med/Surg Advanta 2 (30 total)

Total: 35

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3420-015



Description: Bed, Electric, Birthing

Manufacturer: Hill-Rom - Bed & Stretcher Group (800-445-3730)

Mfr #: P3700 / AF400

Vendor: Unspecified

Vendor #:

Model: Affinity 4 [AF400]

Birthing bed with Prevention Foam Surface and Lift-Off Foot section. Standard Features include: V-Cut or Straight-Cut mattress option, Removable Foot Section, Battery back-up, Powered Foot Section, Four Wheel Dual Locking Casters, OneStep Dampened Siderail Release, Automatic Nightlight, Lock-out Controls, manual Trend;like positioning, Instant CPR, Instant labor grips, and Automatic Seat Tilt (15 degrees).

General Product Detail

Arch Sig: Yes	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 480
Hz: 60	Amps: 4
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 39.00 in (991 mm)	Left: N/A
Depth: 90.00 in (2286 mm)	Right: N/A
Height: 40.00 in (1016 mm)	Front: N/A
Max Weight: 480 lbs (217.7 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4881-006



Description: Bench, Tech Workstation, 6 ft.

Manufacturer: Teclab ((269) 372-6000)

Vendor: Unspecified

Model: TWS-7000-ESD-0

Mfr #: TWS-7000-ESD-0

Vendor #:

6' tech workstation. Includes two 4-drawer, 15" wide base cabinets, one 36" foot rest, one apron drawer, two apron fillers, two 16" tall shelf supports with 5 grounded outlets (plus illuminated combination rocker switch), one plastic laminate work surface, one shelf of same material, and two back guards. ESD-0 option includes wrist strap, ESD work surface, common point ground, 4 jacks, ESD shelf, grounded cabinets, drawer mats, floor mat, and ground monitor.

General Product Detail

Arch Sig: Yes	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: 15
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 72.00 in (1829 mm)	Left: N/A
Depth: 30.00 in (762 mm)	Right: N/A
Height: 56.25 in (1429 mm)	Front: N/A
Max Weight: 396 lbs (179.6 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Product and Project Item Notes

Specification:

Workstation has 3 - 120V hospital grade cords that need to be plugged into 15 amp outlets. Actual amps & watts dependent on equipment plugged into workstation.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Project	Draft (New)	2	
					Total:	2

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 7283-002

AKRO-MILS Description: Bin, Plastic, Allowance
 Manufacturer: Akro-Mils ((800) 253-2467)
 Vendor: Unspecified
 Model: TBD

Mfr #: TBD
 Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	
Total:					3	

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 7788-038



Description: Board, Patient Information

Manufacturer: VividBoard - a GMi Company (800-543-0550)

Vendor: Unspecified

Mfr #: VBINHMY23*

Vendor #: VBINHMY23*

Model: InHarmony Changeable Glassboard 24'x36'

Glass Patient Room Dry Erase Board with fully changeable sliding information panel. 24"x36" inches with custom graphics. Graphic boards slide in and out for easy changes in clinician information or room reconfiguration. Graphics on the glass and changeable panel are custom designed to the facility to match brand and interior spaces. Non-magnetic glass surface promotes infection prevention with smooth non-porous, easily disinfected materials. 1/4 inch tempered, polished glass writing surface with round stand-off hanging system. Available in both vertical and horizontal applications.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: Yes
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 24.00 in (610 mm)	Left: N/A
Depth: 2.00 in (51 mm)	Right: N/A
Height: 36.00 in (914 mm)	Front: N/A
Max Weight: 20 lbs (9.1 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Specification:

Width and height reflects vertical application. Can also be mounted horizontally.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Total:					2	

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 7788-038



Description: Board, Patient Information
Manufacturer: VividBoard - a GMI Company (800-543-0550)
Vendor: VividBoard - a GMI Company (800-543-0550)
Model: InHarmony Changeable Glassboard 24'x36'

Mfr #: VBINHMY23*
Vendor #: VBINHMY23*

Glass Patient Room Dry Erase Board with fully changeable sliding information panel. 24"x36" inches with custom graphics. Graphic boards slide in and out for easy changes in clinician information or room reconfiguration. Graphics on the glass and changeable panel are custom designed to the facility to match brand and interior spaces. Non-magnetic glass surface promotes infection prevention with smooth non-porous, easily disinfected materials. 1/4 inch tempered, polished glass writing surface with round stand-off hanging system. Available in both vertical and horizontal applications.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: Yes
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 24.00 in (610 mm)	Left: N/A
Depth: 2.00 in (51 mm)	Right: N/A
Height: 36.00 in (914 mm)	Front: N/A
Max Weight: 20 lbs (9.1 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Specification:

Width and height reflects vertical application. Can also be mounted horizontally.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	

Total: 4

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5873-042



Description: Board, White, Dry Erase

Manufacturer: Peter Pepper Products, Inc. ((800) 496-0204)

Mfr #: TC7236

Vendor: Unspecified

Vendor #:

Model: Tactics Communication Board 72X36

Dry erase communication Board. 72in.w x 36in.h. Features White porcelain enamel wipe-off writing surface, 28 gauge steel construction with porcelain enamel color cover-coat and non-porous non-absorbent and acid resistant surface. Includes (1) 12 inch aluminum pen rail; (4) dry-erase marking pens; (1) eraser. Wall mounting cleat/bracket & hardware included. 9 frame choices. White 2in. x 2in. Grid Graphics option at up charge. Budget price is for mid grade frames 8-9.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 72.00 in (1829 mm)	Left: N/A
Depth: 2.00 in (51 mm)	Right: N/A
Height: 36.00 in (914 mm)	Front: N/A
Max Weight: 67 lbs (30.4 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	12	Placeholder for supply storage cabinet-mtd. dry erase surfaceThis is only a PLACEHOLDER to budget for the cost to customize the Solaire supply cart #5847-143 with a white board surface
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	3	Placeholder for storage cabinet-mtd. dry erase surfaceThis is only a PLACEHOLDER to budget for the cost to customize the Solaire supply cart #5847-143 with a white board surface
					Total:	15

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3449-000

Description: Bracket, Computer Workstation, Rail
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 27.00 in (686 mm)	Left: N/A
Depth: 43.00 in (1092 mm)	Right: N/A
Height: 34.00 in (864 mm)	Front: N/A
Max Weight: 47 lbs (21.3 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Y	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	8	two computers per bed - mannequin vitals/internal
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	2	two computers per bed - mannequin vitals/internal
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	2	two computers per bed - mannequin vitals/internal
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	60	two computers per bed - mannequin vitals/internal
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	10	two computers per bed - mannequin vitals/internal

Total: 82

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3449-000

Description: Bracket, Computer Workstation, Wall
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 27.00 in (686 mm)	Left: N/A
Depth: 43.00 in (1092 mm)	Right: N/A
Height: 34.00 in (864 mm)	Front: N/A
Max Weight: 47 lbs (21.3 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Y	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	0	wall mounted computer for future inventory management
Storage	467	SIM Equipment Storage	Project	Draft (New)	0	wall mounted for inventory management

Total: 0

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4927-026



Description: Bracket, Patient Transfer Device, Wall Mount

Manufacturer: Techno-Aide ((800) 251-2629)

Vendor: Unspecified

Model: Wall Storage Rack TBR-2P w/ TBS-2372 (23x72 Board)

Mfr #: TBS-2372 / TBR-2P

Vendor #:

Wall mounted patient transfer device storage bracket with 23 x 72 patient transfer board. Semi-flexible, radiolucent, high density polyethylene board. Rack includes two angled 6 inch pegs, white enamel steel. Not anti-static.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 72.00 in (1829 mm)	Left: N/A
Depth: 9.00 in (229 mm)	Right: N/A
Height: 23.00 in (584 mm)	Front: N/A
Max Weight: 15 lbs (6.8 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Wall mounted storage rack mounting holes are on vertical centers. Dimensions reflect board hanging on wall mounted rack.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	New bracket ONLY for 5 existing patient transfer boards
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 2

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3455-065



Description: Cabinet, Patient Room, Bedside
Manufacturer: Stryker Medical ((269) 329-2100)
Vendor: Unspecified
Model: Contemporary MedSurg - 1 Dwr/1 Dr

Mfr #: 4400-800-000
Vendor #: 4400-800-000

Bedside cabinet with 1 drawer and 1 door. Contemporary design with concealed 1.5" (4 cm) dual-wheel casters, finished edges, removable plastic drawers and drawer liners, vented back, steel drawer frames and glides, Thermoform drawer fronts and top, stainless steel pull handles.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 5-Furniture	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 20.00 in (508 mm)	Left: N/A
Depth: 20.00 in (508 mm)	Right: N/A
Height: 30.50 in (775 mm)	Front: N/A
Max Weight: 80 lbs (36.3 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

Total: 41

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 8825-000

Description: Cane, Quad Base
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 4-Instruments	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 4.00 in (102 mm)	Left: N/A
Depth: 6.00 in (152 mm)	Right: N/A
Height: 38.25 in (972 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	12 NM	
Storage	467	SIM Equipment Storage	Project	Draft (New)	2 2 per PF Room NM	

Total: 14

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3581-045



Description: Cart / Trolley, Mannequin, Bulk

Manufacturer: MODRoto Inc. ((800) 829-4535)

Vendor: Unspecified

Model: 101P (46.5 cu.ft.)

Mfr #: 101P

Vendor #:

Bulk linen truck. 60.5in.W x 24in.D x 71.25in.H. Features black powder coated top brace and steel base, torque resistant molded-in inserts, 6 inch rubber casters, drain holes in bottom, hand holes in right side, two plastic shelves (hinged non removable). 800-lb. weight capacity.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 60.50 in (1537 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 71.25 in (1810 mm)	Front: N/A
Max Weight: 165 lbs (74.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	Linen Cart/Trolley & Trolley Cover for Sim-Mannequin storage

Total: 3

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3581-045



Description: Cart / Truck, Linen, Bulk

Manufacturer: MODRoto Inc. ((800) 829-4535)

Vendor: Unspecified

Model: 101P (46.5 cu.ft.)

Mfr #: 101P

Vendor #:

Bulk linen truck. 60.5in.W x 24in.D x 71.25in.H. Features black powder coated top brace and steel base, torque resistant molded-in inserts, 6 inch rubber casters, drain holes in bottom, hand holes in right side, two plastic shelves (hinged non removable). 800-lb. weight capacity.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 60.50 in (1537 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 71.25 in (1810 mm)	Front: N/A
Max Weight: 165 lbs (74.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	Linen Cart/Trolley & Trolley Cover for Sim-Mannequin storage

Total: 1

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5804-116



Description: Cart, Computer, Workstation

Manufacturer: Rubbermaid Healthcare, A Division of Capsa Solutions (888-859-8294)

Mfr #: 1781127

Vendor: Unspecified

Vendor #: 1781127

Model: M38e Computer Cart (40 Amp AC)

Mobile computer cart with power. Features 40 amp Lithium Ion (LiFePO4) battery and direct AC power, internal storage for small form CPU, and a fully adjustable keyboard tray and light with auto shut-off feature. Features contoured handles, document cover, 16" adjustable height work surface, T-Track for universal mounting point for accessories, internal cable management, and CPU storage. Accommodates barcode scanner heads up to 4" x 3.75". 4" precision bearing casters - 2 locking. Maximum LCD monitor weight capacity: 25 lbs. Maximum internal CPU dimensions: 15.25"W x 13.25"D x 3.25"H; maximum keyboard size: 8.75"D x 18.5"W. Work surface size is 21"W x 14"D. Three available power outlets, hospital grade plug, and 2.5 ft. cord extendable to 8 ft.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: REM
Hz: 60	Amps: 20
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 21.00 in (533 mm)	Left: N/A
Depth: 17.00 in (432 mm)	Right: N/A
Height: 61.00 in (1549 mm)	Front: N/A
Max Weight: 101 lbs (45.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Work surface adjusts from 33" to 49"H. Height is to top of monitor (another 12").

Electrical information provided for AC power. 40 amp LiFePO4 battery requires 20 amp circuit for charging.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	6	2 per S&T Room Users requested Vendor product showcase per 4/19/18 mtg.
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5802-030



Description: Cart, Equipment, Fetal Monitor
Manufacturer: GCX Corporation ((707) 773-1100)
Vendor: Unspecified
Model: CO-0025-62 (For Corometrics 250cx)

Mfr #: CO-0025-62
Vendor #:

Fetal monitoring cart. For GE Corometrics 250cx Series Fetal Monitor. Features: 40"/101.6 cm height column, 18"/45.7 cm x 18"/45.7 cm base, 5"/12.7 cm diameter casters (4 locking) for transition over uneven surfaces, Corometrics 250cx Series Mount with paper catcher, four (4) quick release baskets and drawer. Additional drawer for extra storage space, dual grip handle and cable management.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 18.00 in (457 mm)	Left: N/A
Depth: 18.00 in (457 mm)	Right: N/A
Height: 48.00 in (1219 mm)	Front: N/A
Max Weight: 150 lbs (68.0 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	cart for fetal monitor - use this to designate fetal monitor placement
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	cart for fetal monitor - use this to designate fetal monitor placement
Total:					2	

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet



Description: Cart, Medication, w/Computer
Manufacturer: Omnicell, Inc. ((800) 850-6664)
Vendor: Unspecified
Model: Savvy Mobile Cart

Atta ID: 7099-007

Mfr #: MMM-FRM-001

Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 115	Watts: 288
Hz: 60	Amps: 2.5
Phase: Single	BTU/hr: 443
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 17.00 in (432 mm)	Left: N/A
Depth: 22.00 in (559 mm)	Right: N/A
Height: 64.00 in (1626 mm)	Front: 10.00 in (254 mm)
Max Weight: 165 lbs (74.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Max Weight reflects empty with 6" drawers, weight will vary depending on configuration.

Front clearance reflects keyboard fully extended.

Electrical requirements reflect external battery charger. Ensure space in storage area for external battery charger (16.75in x 11in x 7.75in).

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	2	2 designated for the entire PF area This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK

Total: 6

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 8300-009

InnerSpace Description: Cart, Procedure, Accesories, Allowance
 Manufacturer: InnerSpace Corporation (800) 467-7224
 Vendor: Unspecified
 Model: TBD

Mfr #: TBD
 Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 0-Unassigned	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	Point-of-Use storage above-surface extension (to use in absence of a worksurface)
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	Point-of-Use storage above-surface extension (to use in absence of a worksurface)

Total: 14

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5863-324



Description: Cart, Procedure, General

Manufacturer: InnerSpace Corporation ((800) 467-7224)

Vendor: Unspecified

Model: 7300-27W-01-6 White w/Key Lock

Mfr #: 7300-27W-01-6

Vendor #: 7300-27W-01-6

Small procedure cart. White. Features integrated seams, smooth exposed surfaces, integrated handles, scalloped base, left or right pull-out writing surface, 5 inch plate casters (2 locking), key lock. Full extension, self closing drawers: 4-3in., 1-6in., 1-9in. Central locking system with key lock is standard on all models. Optional electronic keyless entry or proximity card access.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 30.25 in (768 mm)	Left: N/A
Depth: 22.25 in (565 mm)	Right: N/A
Height: 39.00 in (991 mm)	Front: N/A
Max Weight: 147 lbs (66.7 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	

Total: 14

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5853-066



Description: Cart, Procedure, Isolation

Manufacturer: InnerSpace Corporation ((800) 467-7224)

Vendor: Unspecified

Model: 7300-27Y-01-5 Isolation Cart

Mfr #: 7300-27Y-01-5

Vendor #: 7300-27Y-01-5

Isolation cart. Yellow. Features integrated seams, smooth exposed surface, integrated handles, key lock, scalloped base, left or right pull-out writing surface, 5 inch plate casters (2 locking). Full extension, self closing drawers: 1-3in. and 4-6in. Central locking system with key lock is standard on all models. Optional electronic keyless entry or proximity card access.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 30.25 in (768 mm)	Left: N/A
Depth: 22.25 in (565 mm)	Right: N/A
Height: 39.00 in (991 mm)	Front: N/A
Max Weight: 147 lbs (66.7 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	
Total:					4	

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BQA043G

Description: Cart, Procedure, Isolation
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #: ST34YN
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 29.00 in (737 mm)	Left: N/A
Depth: 24.50 in (622 mm)	Right: N/A
Height: 45.00 in (1143 mm)	Front: N/A
Max Weight: 99 lbs (44.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5859-085



Description: Cart, Procedure, Resuscitation

Manufacturer: InterMetro Industries Corp (800.992.1776)

Vendor: InterMetro Industries Corp (800.992.1776)

Model: Lifeline LECCRP2

Mfr #: LECCRP2

Vendor #: LECCRP2

Resuscitation Cart with 21" of drawer space and Microban Antimicrobial protection. (1) 3" and (3) 6" drawers. Passive breakaway locking. Includes recessed top storage, wastebasket 28 Quart & Holder, lockable sharps container, adjustable defibrillator tray, O2 tank holder, backboard w/ holder, IV pole and 100 security seals.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: Yes
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 38.00 in (965 mm)	Left: N/A
Depth: 22.75 in (578 mm)	Right: N/A
Height: 38.75 in (984 mm)	Front: N/A
Max Weight: 131 lbs (59.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	227	Alcove, Code Cart	Project	Draft (New)	2	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Existing (Reuse)	Draft (New)	1	

Total: 6

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5859-088



Description: Cart, Procedure, Resuscitation

Manufacturer: InterMetro Industries Corp (800.992.1776)

Vendor: Unspecified

Model: Lifeline LECP13000

Mfr #: LECP13000

Vendor #:

Resuscitation Cart with 21" of drawer space and Microban Antimicrobial protection. Configured with one 3" and three 6" drawers, tank holder, 5th wheel steering assist, Code Blue drawer pull color.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: Yes
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 38.00 in (965 mm)	Left: N/A
Depth: 22.75 in (578 mm)	Right: N/A
Height: 38.25 in (972 mm)	Front: N/A
Max Weight: 131 lbs (59.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 1

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5860-013



Description: Cart, Procedure, Resuscitation, Pediatric

Manufacturer: InterMetro Industries Corp (800.992.1776)

Vendor: Unspecified

Model: Lifeline LECPEDS2

Mfr #: LECPEDS2

Vendor #:

Resuscitation Cart with 27" of drawer space and Microban Antimicrobial protection. (7) 3" and (1) 6" drawers. Includes (7) 3" and (1) 6" Drawer Divider Tray with Dividers, (4) locking side bins, adjustable defibrillator tray, lockable sharps container, cord manager, Hospital Grade 6-Outlet Strip and Holder, O2 tank holder, suction pump shelf, Backboard with Back Assembly Kit, IV pole, individual drawer seal lock bar, 100 security seals, drawer dividers and pediatric code response drawer kit.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: Yes
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 38.00 in (965 mm)	Left: N/A
Depth: 22.75 in (578 mm)	Right: N/A
Height: 45.50 in (1156 mm)	Front: N/A
Max Weight: 177 lbs (80.3 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	227	Alcove, Code Cart	Project	Draft (New)	1	
Total:					1	

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ870V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	Allowance for Dbl Dr Supply Cart accesories

Total: 1

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ872V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	Allowance for Dbl Dr Supply Cart accesories

Total: 4

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ873V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	Allowance for Dbl Dr Supply Cart accesories

Total: 1

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ874V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	Allowance for Dbl Dr Supply Cart accesories

Total: 1

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ875V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	Allowance for Dbl Dr Supply Cart accesories

Total: 10

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: BRJ876V

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	Allowance for Dbl Dr Supply Cart accesories

Total: 4

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: CS550SM

Description: Cart, Supply, Accesories, Allowance
Manufacturer: Solaire Medical
Vendor: Unspecified
Model: Unspecified

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	Allowance for Dbl Dr Supply Cart accesories

Total: 3

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5857-143



Description: Cart, Supply, Dbl Dr

Manufacturer: Solaire Medical (888-435-2256)

Vendor: Solaire Medical (888-435-2256)

Model: Roam 1.0 Glass Door Double 3210GVC

Mfr #: 3210GVC

Vendor #: 3210GVC

Enclosed supply cart. Hinged double glass doors with aluminum extrusions. 28.75D x 40.25W x 80H. Features aluminum construction, push/pull stainless steel handle, removable center column, 5" medical grade plate casters, central key lock and metal ceiling inside pre-set to hold up to seven optional Cath Cues. 58 Usable Flexcells, 29 per side. Cart can be optimized for product specific storage and side accessories. Accessories are sold separately. Electronic Lock option available. For interior accessories, add approximately \$1,300.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 40.50 in (1029 mm)	Left: N/A
Depth: 28.75 in (730 mm)	Right: N/A
Height: 80.00 in (2032 mm)	Front: N/A
Max Weight: 197 lbs (89.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	12	This cart to be customized with a white board surface on the back
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	3	This cart to be customized with a white board surface on the back
Total:					15	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6339-004



Description: Cart, Supply, Linen
Manufacturer: Lakeside Manufacturing, Inc. ((800) 558-8565)
Vendor: Unspecified
Model: 332 w/cover

Mfr #: 332
Vendor #:

Linen supply cart. Standard duty, 300 lbs capacity. Features all-welded stainless steel construction, four 15.5in. x 24in. shelves with 11 3/4 inch clearance, 5 inch casters (2 swivel, 2 fixed) & bumpers on legs and handles. Includes nylon cart cover. 27.50in.W x 16.25in.D x 46in.H.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 27.50 in (699 mm)	Left: N/A
Depth: 16.25 in (413 mm)	Right: N/A
Height: 46.00 in (1168 mm)	Front: N/A
Max Weight: 36 lbs (16.3 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	2	Moved here from Skills and Tasks. It will live here and be utilized as a linen exchanged cart as needed - per meeting 3/16/18

Total: 2

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5857-146



Description: Cart, Supply, Single Dr

Manufacturer: Solaire Medical (888-435-2256)

Vendor: Unspecified

Model: Roam 0.25 Glass Door Single S3225RGVC

Mfr #: S3225RGVC

Vendor #: S3225RGVC

Enclosed supply cart. Single wide with right hinged glass door. 24.25"D x 29"W x 70"H. Features aluminum construction, push/pull stainless steel handle, 5" medical grade plate casters, central key lock and metal ceiling inside pre-set to hold up to five optional Cath Cues. 25 Usable Flexcells. Cart can be optimized for product specific storage and side accessories. Accessories are sold separately. Electronic Lock option available.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 29.00 in (737 mm)	Left: N/A
Depth: 24.25 in (616 mm)	Right: N/A
Height: 70.00 in (1778 mm)	Front: N/A
Max Weight: 130 lbs (59.0 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	
					Total:	20

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5863-330



Description: Cart, Task Trainers

Manufacturer: InnerSpace Corporation ((800) 467-7224)

Vendor: Unspecified

Model: 7300-33W-01-7 White w/Key Lock

Mfr #: 7300-33W-01-7

Vendor #:

Small procedure cart. White. Features integrated seams, smooth exposed surfaces, integrated handles, scalloped base, left or right pull-out writing surface, 5 inch plate casters (2 locking), key lock. Full extension, self closing drawers: 4-3in., 2-6in., 1-9in. Central locking system with key lock is standard on all models. Optional electronic keyless entry or proximity card access.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 30.25 in (768 mm)	Left: N/A
Depth: 22.25 in (565 mm)	Right: N/A
Height: 44.75 in (1137 mm)	Front: N/A
Max Weight: 166 lbs (75.3 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	7	
Total:					7	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5836-015



Description: Cart, Utility, Polymer
Manufacturer: Rubbermaid Commercial Products (800-347-9800)
Vendor: Unspecified
Model: 4091 Xtra (Black)

Mfr #: FG409100BLA
Vendor #: FG409100BLA

Open sided utility cart. Three large black shelves with push-handles on both sides of cart. 300 lb. capacity.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 20.00 in (508 mm)	Left: N/A
Depth: 40.75 in (1035 mm)	Right: N/A
Height: 38.00 in (965 mm)	Front: N/A
Max Weight: 43 lbs (19.5 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Project	Draft (New)	5	
Storage	467	SIM Equipment Storage	Project	Draft (New)	3	
Total:					8	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3603-029



Description: Chair, Clinical, Commode, Floor

Manufacturer: Medline Industries Inc. ((847) 949-5500)

Vendor: Unspecified

Model: Foldable Commode

Mfr #: MDS89664F

Vendor #:

Foldable Commode Chair. Features aluminum frame, plastic seat with lid, plastic armrests, adjustable legs, seat height adjustment 16" to 22" with 1" increments, rubber tips, 10 quart commode bucket with cover and splash shield. 300 lb weight capacity. Ships 4/case.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 22.00 in (559 mm)	Left: N/A
Depth: 16.50 in (419 mm)	Right: N/A
Height: 27.00 in (686 mm)	Front: N/A
Max Weight: 15 lbs (6.8 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Existing (Reuse)	Draft (New)	3	
Storage	467	SIM Equipment Storage	Existing (Reuse)	Draft (New)	1	
Total:					4	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5177-020



Description: Clock, Analog, Wall

Manufacturer: Chicago Lighthouse Industries ((312) 666-1331)

Mfr #: 67300040

Vendor: Unspecified

Vendor #:

Model: 24 Hour Series [67300040]

Wall mounted, battery operated clock. 12 and 24-hour dial rings, numerals and minute marks, red sweep second hand.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 12.75 in (324 mm)	Left: N/A
Depth: 2.00 in (51 mm)	Right: N/A
Height: 12.75 in (324 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Requires 1 AA battery.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Administration	484	Office, Director - Administration	Project	Draft (New)	1	
Administration	482	Office, General - Administration	Project	Draft (New)	1	
Administration	480	Office, Tech	Project	Draft (New)	1	
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 6

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6978-017



Description: Clock, Digital, Synchronized, Wireless

Manufacturer: Innovation Wireless ((888) 559-5565)

Vendor: Unspecified

Model: Digital Display Clock 642307

Mfr #: 642307

Vendor #: 642307

Wireless synchronized digital clock with red LED 4 inch - 4 digit, single sided display. 12/24 hour format, adjustable brightness. Wireless connectivity to KRONOsync transmitter (sold separately).

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 12
Hz: 60	Amps: 0.1
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 19.75 in (502 mm)	Left: N/A
Depth: 2.50 in (64 mm)	Right: N/A
Height: 6.50 in (165 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470K	Control Room	Project	Draft (New)	0	
Patient Care Flex	470D, 470,G, 470H, 475	Debrief Room	Project	Draft (New)	0	
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	0	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	0	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	0	

Total: 0

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4988-000

Description: Computer, Desktop
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: Yes

Electrical Requirements

Volts: 120	Watts: 280
Hz: 60	Amps: 2.3
Phase: Single	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 6.75 in (171 mm)	Left: N/A
Depth: 16.75 in (425 mm)	Right: N/A
Height: 15.50 in (394 mm)	Front: N/A
Max Weight: 21 lbs (9.5 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Administration	484	Office, Director - Administration	Project	Draft (New)	0	
Administration	482	Office, General - Administration	Project	Draft (New)	0	
Administration	480	Office, Tech	Project	Draft (New)	0	
Patient Care Flex	470D, 470,G, 470H, 475	Debrief Room	Project	Draft (New)	0	
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	0	

Total: 0

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 5025-002



Description: Cover, Cart, Allowance
Manufacturer: InterMetro Industries Corp (800.992.1776)
Vendor: Unspecified
Model: TBD

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	Trolley Cover for Sim-Mannequin storage
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	Cart Cover for Linen Cart that hold the SIM MannequinsLinen Cart/Trolley & Trolley Cover for Sim-Mannequin storage

Total: 4

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 8821-000

Description: Crutches, Axillary, Adult
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 4-Instruments	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	12 NM	
Storage	467	SIM Equipment Storage	Project	Draft (New)	2 NM	2 per PF Room

Total: 14

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3678-047



Description: Defibrillator, Monitor, w/Pacing

Manufacturer: Zoll Medical Corporation ((978) 421-9655)

Vendor: Zoll Medical Corporation ((978) 421-9655)

Model: R Series ALS w/Exp Pkg, Masimo SpO2, ETCO2, NIBP

Mfr #: 3 0720 0052 0131 0012

Vendor #: 3 0720 0052 0131 0012

Defibrillator with Pacing, Masimo Pulse Oximetry, ETCO2, NIBP and Expansion package (Real CPR Help which offers quantitative depth and rate in real time, release indicator, interruption timer, See-Thru CPR ECG artifact filtering, WiFi compatibility). Includes audible and visual advisory prompts, single cable capable of monitoring/ pacing/defib, code readiness testing system, rectilinear biphasic technology, compact flash card slot, advisory mode, integrated AC power, built-in test port, 5-year warranty. Rechargeable Lithium Ion Battery is NOT included with device and must be ordered as a separate line item. Order part # 8019-0535-01 at a price of \$475.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: Yes	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 10.50 in (267 mm)	Left: N/A
Depth: 12.50 in (318 mm)	Right: N/A
Height: 8.25 in (210 mm)	Front: N/A
Max Weight: 15 lbs (6.8 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	227	Alcove, Code Cart	Project	Draft (New)	3	on crash cart

Total: 3

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 8794-002



Description: Diagnostic System, Integrated

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company ((800) 535-6663)

Mfr #: 77791-2MPX

Vendor: Unspecified

Vendor #:

Model: Green Series 777 [77791-2MPX]

Integrated Diagnostic System. Includes Wall Aneroid Sphygmomanometer (#7670-01), SureTemp Plus 690 Electronic Thermometer (#01690-300), PanOptic Ophthalmoscope with Cobalt-Blue Filter and Corneal Viewing Lens (#11820), MacroView Otoscope (#23810), KleenSpec Disposable Specula Dispenser, Wall-Mount Panel (30 x 12 in). Wall Transformer handles automatically turn on/off. 12 ft. cord length. Energy Star Qualified.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: Yes

Electrical Requirements

Volts: 120	Watts: 22
Hz: 60	Amps: 0.2
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 30.00 in (762 mm)	Left: N/A
Depth: 6.00 in (152 mm)	Right: N/A
Height: 12.00 in (305 mm)	Front: N/A
Max Weight: 16 lbs (7.3 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Product and Project Item Notes

Specification:

Dimensions reflect wall board to which diagnostic products are mounted. See cutsheet for details.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	ENT wall system for headwall
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	ENT wall system for headwall

Total: 14

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 7050-025



Description: Dispenser, Glove, Quadruple Box

Manufacturer: Health Care Logistics (614-853-6798)

Vendor: Unspecified

Model: 17899 Stainless Steel

Mfr #: 17899

Vendor #:

Quadruple stainless steel glove box holder with dividers. Keyholes for mounting vertically or horizontally.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 21.00 in (533 mm)	Left: N/A
Depth: 3.50 in (89 mm)	Right: N/A
Height: 10.00 in (254 mm)	Front: N/A
Max Weight: 4 lbs (1.8 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Keyholes for mounting vertically or horizontally.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	1	
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 5869-011



Description: Dispenser, Hand Sanitizer, Wall Mount
Manufacturer: GOJO Industries (800.321.9647)
Vendor: Unspecified
Model: Purell APX Aerosol (9699-12)

Mfr #: 9699-12
Vendor #: 9699-12

Wall mounted hand sanitizer dispenser. For use with Purell hand sanitizer. Push-bar actuation. Fully ADA compliant. Ships 12/case.
**Can be purchased through Distributor or provided free-of-charge with purchase of refills.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 3.25 in (83 mm)	Left: N/A
Depth: 4.25 in (108 mm)	Right: N/A
Height: 4.50 in (114 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470K	Control Room	Lease	Draft (New)	1	existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
Patient Care Flex	470D, 470,G, 470H, 475	Debrief Room	Lease	Draft (New)	4	existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
Patient Care Flex	470E	Nursing Alcove	Lease	Draft (New)	1	existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
Patient Care Flex	470A	Patient Flex, Maternity / OB	Lease	Draft (New)	1	1 designated for the entire PF Room Existing wall mtd. sanitizer dispenser. actual sanitizer typically purchased via distributor agreement
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Lease	Draft (New)	6	1 @ sink, 1 @ door
Skills and Tasks	460	Skills & Tasks, Bed (5)	Lease	Draft (New)	2	1 @ sink, 1 @ door
Standardized Patients	401 - 421 (Odd	Exam Room	Lease	Draft (New)	10	

	numbers)				
Standardized Patients	423, 425, 427, 429	Psy Consultation	Lease	Draft (New)	4
				Total:	29

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5869-011



Description: Dispenser, Hand Sanitizer, Wall Mount

Manufacturer: GOJO Industries (800.321.9647)

Vendor: GOJO Industries (800.321.9647)

Model: Purell APX Aerosol (9699-12)

Mfr #: 9699-12

Vendor #: 9699-12

Wall mounted hand sanitizer dispenser. For use with Purell hand sanitizer. Push-bar actuation. Fully ADA compliant. Ships 12/case.

**Can be purchased through Distributor or provided free-of-charge with purchase of refills.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 3.25 in (83 mm)	Left: N/A
Depth: 4.25 in (108 mm)	Right: N/A
Height: 4.50 in (114 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6084-061



Description: Dispenser, Paper Towel, Surface Mount

Manufacturer: SCA Tissue North America ((866) 722-8675)

Vendor: Unspecified

Model: Tork Elevation Matic (5510282)

Mfr #: 5510282

Vendor #: 5510282

Surface mounted paper towel dispenser. Hands-free dispensing, dispenses single towel at a time when one pulled. Push bar for loading and emergency feeds. Uses any Tork H1 roll towel.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 13.25 in (337 mm)	Left: N/A
Depth: 8.25 in (210 mm)	Right: N/A
Height: 14.75 in (375 mm)	Front: N/A
Max Weight: 7 lbs (3.2 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Lease	Draft (New)	1	Existing paper towel dispenser
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Lease	Draft (New)	4	Existing paper towel dispenser
Patient Care Flex	470A	Patient Flex, Maternity / OB	Lease	Draft (New)	1	Existing paper towel dispenser
Patient Care Flex	470L	Patient Flex, NICU	Lease	Draft (New)	1	Existing paper towel dispenser
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Lease	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Lease	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Lease	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Lease	Draft (New)	4	

Total: 25

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 6084-061



Description: Dispenser, Paper Towel, Surface Mount

Manufacturer: SCA Tissue North America ((866) 722-8675)

Mfr #: 5510282

Vendor: SCA Tissue North America ((866) 722-8675)

Vendor #: 5510282

Model: Tork Elevation Matic (5510282)

Surface mounted paper towel dispenser. Hands-free dispensing, dispenses single towel at a time when one pulled. Push bar for loading and emergency feeds. Uses any Tork H1 roll towel.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 13.25 in (337 mm)	Left: N/A
Depth: 8.25 in (210 mm)	Right: N/A
Height: 14.75 in (375 mm)	Front: N/A
Max Weight: 7 lbs (3.2 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5868-069



Description: Dispenser, Soap, Wall Mount
Manufacturer: GOJO Industries (800.321.9647)
Vendor: Unspecified
Model: GOJO FMX-12 Black (5155-06)

Mfr #: 5155-06
Vendor #: 5155-06

Wall mounted soap dispenser. Black. Holds 1250 mL sealed refill. Skylight with extra-large site window. ADA compliant for one hand push operation. Optional key lock. Ships 6/case.
 **Can be purchased through Distributor or provided free-of-charge with purchase of refills.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 6.25 in (159 mm)	Left: N/A
Depth: 5.25 in (133 mm)	Right: N/A
Height: 10.00 in (254 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:
 Mounts with adhesive strips (included) or screws (not included).

Ships 6/case.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Lease	Draft (New)	1	Existing dispenser
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Lease	Draft (New)	4	Existing dispenser
Patient Care Flex	470A	Patient Flex, Maternity / OB	Lease	Draft (New)	1	Existing dispenser
Patient Care Flex	470L	Patient Flex, NICU	Lease	Draft (New)	1	Existing dispenser
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Lease	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Lease	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Lease	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Lease	Draft (New)	4	

Total: 25

UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 5868-069



Description: Dispenser, Soap, Wall Mount

Manufacturer: GOJO Industries (800.321.9647)

Vendor: GOJO Industries (800.321.9647)

Model: GOJO FMX-12 Black (5155-06)

Mfr #: 5155-06

Vendor #: 5155-06

Wall mounted soap dispenser. Black. Holds 1250 mL sealed refill. Skylight with extra-large site window. ADA compliant for one hand push operation. Optional key lock. Ships 6/case.

**Can be purchased through Distributor or provided free-of-charge with purchase of refills.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 6.25 in (159 mm)	Left: N/A
Depth: 5.25 in (133 mm)	Right: N/A
Height: 10.00 in (254 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Mounts with adhesive strips (included) or screws (not included).

Ships 6/case.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Lease	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3723-039



Description: Disposal, Sharps, Wall Mount
Manufacturer: BD Medical ((201) 847-6800)
Vendor: Unspecified
Model: Next Generation 305435 (2 Gal) w/ 305409 Bracket

Mfr #: 305435/305409
Vendor #: 305435/305409

Wall mounted, locking bracket with red 2 gallon sharps container with counterbalanced door and horizontal entry. Sharps container ships 10/case, Bracket ships 10/case.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 14.00 in (356 mm)	Left: N/A
Depth: 4.50 in (114 mm)	Right: N/A
Height: 6.00 in (152 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gases: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Dimensions reflect bracket with sharps container.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Lease	Draft (New)	4	Existing sharps 2 gallon
Patient Care Flex	470A	Patient Flex, Maternity / OB	Lease	Draft (New)	1	Existing sharps 2 gallon
Patient Care Flex	470L	Patient Flex, NICU	Lease	Draft (New)	1	Existing sharps 2 gallon
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Lease	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Lease	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Lease	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Lease	Draft (New)	4	

Total: 24

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3723-039



Description: Disposal, Sharps, Wall Mount

Manufacturer: BD Medical ((201) 847-6800)

Vendor: BD Medical ((201) 847-6800)

Model: Next Generation 305435 (2 Gal) w/ 305409 Bracket

Mfr #: 305435/305409

Vendor #: 305435/305409

Wall mounted, locking bracket with red 2 gallon sharps container with counterbalanced door and horizontal entry. Sharps container ships 10/case, Bracket ships 10/case.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 14.00 in (356 mm)	Left: N/A
Depth: 4.50 in (114 mm)	Right: N/A
Height: 6.00 in (152 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Dimensions reflect bracket with sharps container.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Lease	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6145-047



Description: Doppler, Vascular

Manufacturer: Natus Neurology - Nicolet Dopplers ((608) 829-8500)

Mfr #: ED25

Vendor: Unspecified

Vendor #:

Model: Elite 200 (2 & 5 MHz probes)

Vascular doppler with 2 MHz obstetric and 5 MHz vascular probes. Features digital display, built-in speaker, autocorrelation, and uses a 9V alkaline battery. Probes can be gas sterilized or sheathed.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 3.00 in (76 mm)	Left: N/A
Depth: 6.75 in (171 mm)	Right: N/A
Height: 1.00 in (25 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Requires 9V battery.
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Existing (Reuse)	Draft (New)	3	NM
Skills and Tasks	460	Skills & Tasks, Bed (5)	Existing (Reuse)	Draft (New)	1	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	NM

Total: 5

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3768-079



Description: Electrocardiograph (ECG), Interpretive
Manufacturer: Mortara Instrument, Inc. (800.231.7437)
Vendor: Unspecified
Model: Burdick ELI 250c Interpretive w/ Stand (USB)

Mfr #: BUR250C-81X / 9911-016-50
Vendor #:

Interpretive ECG, 12-lead with cart. Features backlit LCD display, full keyboard with one-touch buttons for ECG acquisition, printing, and transmission. Full-size paper output includes multiple print options including 3+1, 3+3, 6, 6 + 6, or 12-channel outputs. Internal storage for up to 40 ECGs. Patient demographics can be exchanged with a variety of third-party systems. Includes: AM12 Patient cable, USB / LAN 802.11 connections and power cord. For additional configurations, please contact Mortara Instrument.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 110	Watts: 110
Hz: 60	Amps: 1
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 20.00 in (508 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 55.00 in (1397 mm)	Front: N/A
Max Weight: 52 lbs (23.6 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Internally rechargeable battery.
Weight and dimensions include stand.
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 1

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 3806-000

Description: Flowmeter, Air
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 1.25 in (32 mm)	Left: N/A
Depth: 3.50 in (89 mm)	Right: N/A
Height: 5.50 in (140 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	on headwall
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	on headwall
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	on headwall
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	on headwall
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

Total: 41

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 3803-000

Description: Flowmeter, Oxygen
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 1.25 in (32 mm)	Left: N/A
Depth: 3.50 in (89 mm)	Right: N/A
Height: 5.50 in (140 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	on headwall
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	on headwall
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	on headwall
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	on headwall
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

Total: 41

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 3836-114



Description: Hamper, Linen

Manufacturer: Blickman Inc. ((800) 247-5070)

Vendor: Unspecified

Model: 2010A Chrome (Assembled)

Mfr #: 0962010100

Vendor #: 0962010100

Step-on, chrome linen hamper with white polycoated steel lid. Supports 36-42 gallon multi-purpose bags with a ring to keep bags in place. Includes hamper bag storage pouch (2012). 3" hard wheel casters, two with locks. Hamper bags optional. Ships assembled.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 18.75 in (476 mm)	Left: N/A
Depth: 19.50 in (495 mm)	Right: N/A
Height: 37.75 in (959 mm)	Front: N/A
Max Weight: 25 lbs (11.3 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	1	1 designated for the entire PF Room
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	
Total:					5	

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4860-014



Description: Headwall Column, Floor to Ceiling

Manufacturer: Hospital Systems, Inc. (925.427.7800)

Vendor: Hospital Systems, Inc. (925.427.7800)

Model: Proxima

Mfr #: 1904239

Vendor #: 1904239

Floor-to-Ceiling Patient Service Column for Critical Care areas. Features integral Unimount Accessory Tracks. May be configured with clock timer, medical gas outlets, power outlets, monitor provision, Bedcom Interface, nurse call provision, storage accessories. Custom built: Final weight/dimensions/utilities to be determined. Made in USA. Green certified.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: Yes

Electrical Requirements

Volts: 120	Watts: 2400
Hz: 60	Amps: 20
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: Yes	Plug Type: Hardwire

Physical Requirements

Width: 21.00 in (533 mm)	Left: N/A
Depth: 6.00 in (152 mm)	Right: N/A
Height: REM	Front: N/A
Max Weight: REM	Back: N/A
Mounting: Special	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / Yes
Tech Connect: Yes	

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Product and Project Item Notes

Specification:

Custom built: Final weight/dimensions/utilities to be determined.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	8	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accessories
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accessories
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	15	This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK

Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	3	Headwall with vertical rails for monitors and patient assessment accessories This equipment has an ALTERNATE selection to consider prior to purchase. PLEASE REFER TO MEDICAL EQUIPMENT WORKBOOK Headwall with vertical rails for monitors and patient assessment accessories
					Total:	27

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3843-020



Description: Headwall, Flatwall, 1 Patient

Manufacturer: Hospital Systems, Inc. (925.427.7800)

Mfr #: 1904159

Vendor: Hospital Systems, Inc. (925.427.7800)

Vendor #: 1904159

Model: Axiom Surface Mounted Headwall

Vertical headwall. Includes 1 oxygen, 1 air, 1 vacuum outlet, 5 duplex receptacles, 1 switch for exam or over head light, night light, 1 vacuum bottle slides, 1 data/telephone provision, 1 code blue and 1 nurse call provision. Made in USA. Green Certified.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: Yes

Electrical Requirements

Volts: 120	Watts: REM
Hz: 60	Amps: 20
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: Yes	Plug Type: Hardwire

Physical Requirements

Width: 23.75 in (603 mm)	Left: N/A
Depth: 4.75 in (121 mm)	Right: N/A
Height: 90.00 in (2286 mm)	Front: N/A
Max Weight: REM	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / Yes
Tech Connect: Yes	

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Product and Project Item Notes

Specification:

Dimensions and utility requirements may vary. All HSI headwalls are Build-to-Order and will be assigned their own unique model numbers during the submittal phase.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	8	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	15	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	2	
Total:					26	

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 7249-003



Description: Light, Treatment, Phototherapy, Infant, In-bed

Manufacturer: Natus Newborn Care ((650) 802-0400)

Vendor: Unspecified

Model: neoBLUE Blanket LED Phototherapy System

Mfr #: 006244

Vendor #:

Phototherapy light, in-bed, for infant. Utilizes blue LED technology which emits blue light in the 450 – 475 nm spectrum, matching the peak absorption wavelength (458 nm) at which bilirubin is broken down. Consists of light box and large size, oval blanket. Automatically shuts off in the event of light box overheating; flashing indicator light alerts user to check for blocked air vents. Mattresses and disposable covers are optional and must be ordered separately.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 192
Hz: 60	Amps: 1.6
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 4.75 in (121 mm)	Left: N/A
Depth: 9.25 in (235 mm)	Right: N/A
Height: 5.50 in (140 mm)	Front: N/A
Max Weight: 3 lbs (1.4 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Dimensions shown are for light box.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Total:					1	

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 6403-027



Description: Monitor, Blood Glucose, Point-of-Care

Manufacturer: Roche Diagnostics Corporation (800-428-5076)

Vendor: Unspecified

Model: ACCU-CHEK Performa Meter

Mfr #: 05959551001

Vendor #:

Hand held monitor for blood glucose testing. Results in 5 seconds from 0.6 µ sample, sample detection. Prevents the use of expired test strips. Nonslip rubber grips and large. Uses Performa test strips only. CLIA waived and cleared for multi patient use.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 2.25 in (57 mm)	Left: N/A
Depth: 1.00 in (25 mm)	Right: N/A
Height: 3.75 in (95 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Existing (Reuse)	Draft (New)	15	NM
Skills and Tasks	460	Skills & Tasks, Bed (5)	Existing (Reuse)	Draft (New)	5	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	NM
Total:					21	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5254-017



Description: Monitor, O.B., Intrapartum, Maternal/Fetal

Manufacturer: GE Healthcare - Monitoring Systems (800-558-5120)

Mfr #:

Vendor: GE Healthcare - Monitoring Systems (800-558-5120)

Vendor #:

Model: Corometrics 259cx

Intrapartum Maternal/fetal monitor. Prenatal & postpartum assessments, fetal ECG, TOCO/IUP and independent monitoring for twins. NIBP, maternal SpO2, maternal ECG. Maternal Vital Signs history documents up to eight hours of maternal parameters.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 48
Hz: 60	Amps: 0.4
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 16.50 in (419 mm)	Left: N/A
Depth: 17.50 in (445 mm)	Right: N/A
Height: 6.75 in (171 mm)	Front: N/A
Max Weight: 24 lbs (10.9 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Existing (Reuse)	Draft (New)	1	Existing - sits on cart NM (use cart for footprint)
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	Sits on cart NM (use cart for footprint)

Total: 2

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4071-073



Description: Monitor, Physiologic, Vital Signs, w/Stand

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company ((800) 535-6663)

Mfr #: 68MXTX-B / 4800-60

Vendor: Unspecified

Vendor #:

Model: Connex 6800 Wireless w/ Stand

Vital Signs Monitor with mobile stand. Features: Masimo SpO2, SureTemp Plus Thermometry, SureBP Non-invasive Blood Pressure, Pulse Rate, MAP, BP Averaging, Nurse Call, Four USB Ports for Accessories, Internal 802.11 a/b/g Wireless Radio; 100-240 V, 50-60 Hz AC, IEC Plug Type-B.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 180
Hz: 60	Amps: 1.5
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: Yes	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 21.00 in (533 mm)	Left: N/A
Depth: 21.00 in (533 mm)	Right: N/A
Height: 54.00 in (1372 mm)	Front: N/A
Max Weight: 37 lbs (16.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Can be powered via AC power or battery.

Dimensions include mobile stand.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470	Patient Flex Corridor	Project	Draft (New)	4	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	15	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	3	
Standardized Patients	4H01	SP Corridor - Exam	Project	Draft (New)	1	
Standardized Patients	4H02	SP Corridor - Psy	Project	Draft (New)	1	
Total:					24	

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6852-001

Hill-Rom

Description: Nurse Call, Interface, Bed

Manufacturer: Hill-Rom - Communications ((800) 445-3730)

Vendor: Unspecified

Model: TBD

Mfr #: TBD

Vendor #: TBD

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: C/C	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	on headwall
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	on headwall
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	on headwall

Total: 6

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4091-039



Description: Oto/Ophthalmoscope Set, Rail Mount

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company ((800) 535-6663)

Vendor: Unspecified

Model: Green Series 777 [77510/11710/25020]

Mfr #: 77510/11710/25020

Vendor #: 77510/11710/25020

Wall Transformer #77710 with KleenSpec Disposable Specula Dispenser #52400-PF. Also includes 3.5 V Standard Ophthalmoscope #11710 and 3.5 V Diagnostic Otoscope #25020 with specula. Handles automatically turn on/off. 12 ft. cord length.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 24
Hz: 60	Amps: 0.2
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 9.25 in (235 mm)	Left: N/A
Depth: 5.00 in (127 mm)	Right: N/A
Height: 8.25 in (210 mm)	Front: N/A
Max Weight: 3 lbs (1.4 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	Rail mounted for headwall
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	Rail mounted for headwall
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	Rail mounted for headwall

Total: 6

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4091-040



Description: Oto/Ophthalmoscope Set, Rail Mount

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company ((800) 535-6663)

Vendor: Unspecified

Model: Green Series 777 [77710/11710/25020]

Mfr #: 77710/11710/25020

Vendor #: 77710/11710/25020

Wall mounted oto/ophthalmoscope set. Configuration includes: wall transformer with 3.5V Standard Ophthalmoscope (#11710) with 28 focusing lenses with a range of -25 to +40 diopters. 3.5V Diagnostic Oscope (#25020) with specula provides halogen lamp illumination, wide-angle viewing lens. Handles automatically turn on/off. 12 ft. cord length.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 24
Hz: 60	Amps: 0.2
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 9.25 in (235 mm)	Left: N/A
Depth: 5.00 in (127 mm)	Right: N/A
Height: 8.25 in (210 mm)	Front: N/A
Max Weight: 3 lbs (1.4 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	Rail mounted for both power column and headwall
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	Rail mounted for both power column and headwall

Total: 35

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4109-011



Description: Oximeter, Pulse, Hand Held

Manufacturer: GE Healthcare - Anesthesia Delivery ((800) 345-2700)

Mfr #: 6051-0000-159

Vendor: Unspecified

Vendor #:

Model: TuffSat

Handheld pulse oximeter. Utilizes Relative Perfusion Index Measurement for a dynamic numeric reflection of perfusion at the sensor site. Includes carrying case with belt clip.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O Type: Medical	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 3.00 in (76 mm)	Left: N/A
Depth: 1.25 in (32 mm)	Right: N/A
Height: 6.00 in (152 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Existing (Reuse)	Draft (New)	3	NM
Skills and Tasks	460	Skills & Tasks, Bed (5)	Existing (Reuse)	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Existing (Reuse)	Draft (New)	10	NM
Standardized Patients	423, 425, 427, 429	Psy Consultation	Existing (Reuse)	Draft (New)	4	NM

Total: 18

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6922-001

Description: Printer, Allowance
Manufacturer: Hewlett-Packard (800-474-6836)
Vendor: Unspecified
Model: TBD

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	0	printer for future inventory management
Storage	467	SIM Equipment Storage	Project	Draft (New)	0	printer for future inventory management

Total: 0

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 5866-000

Description: Printer, Laser
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: Yes

Electrical Requirements

Volts: 120	Watts: 495
Hz: 60	Amps: 4.1
Phase: Single	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 14.50 in (368 mm)	Left: N/A
Depth: 14.25 in (362 mm)	Right: N/A
Height: 7.25 in (184 mm)	Front: N/A
Max Weight: 15 lbs (6.8 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Administration	484	Office, Director - Administration	Project	Draft (New)	0	
Administration	482	Office, General - Administration	Project	Draft (New)	0	Existing
Administration	480	Office, Tech	Project	Draft (New)	0	Existing

Total: 0

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6643-003



Description: Pump, Infusion, Controller, Modular
Manufacturer: BD - Alaris Infusion
Vendor: Unspecified
Model: Alaris PC Unit (8015)

Mfr #: 8015
Vendor #:

Central programming and power supply component for the Alaris System infusion modules. Color display, integrated wireless connectivity, Guardrails Suite MX wireless software.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O Type: Medical	Green: No

Electrical Requirements

Volts: 120	Watts: 150
Hz: 60	Amps: 1.3
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 7.00 in (178 mm)	Left: N/A
Depth: 9.00 in (229 mm)	Right: N/A
Height: 9.00 in (229 mm)	Front: N/A
Max Weight: 8 lbs (3.6 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	Plus 4 Modules: PCA, Syringe, (2) Pump Modules - per meeting 3/16/18
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	6	Controller plus one - 2 Modules: (2) Pump Set and one - 1 Module: (1) Pump Module- per Young 4/19/18
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	2	Controller plus one - 2 Modules: (2) Pump Set and one - 1 Module: (1) Pump Module- per Young 4/19/18

Total: 14

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4181-008



Description: Pump, Infusion, PCA
Manufacturer: BD - Alaris Infusion
Vendor: Unspecified
Model: Alaris PCA Pump Module (8120)

Mfr #: 8120
Vendor #:

PCA infusion pump module to be used with Alaris PC unit. Integrates a syringe-based patient-controlled analgesia (PCA) device with large volume pump (LVP). Features flow rate programming range 0.1 to 999mL/hr, rate restriction by syringe size, wireless connectivity. Offers PCA delivery and continuous respiratory monitoring on one platform. The PCA pause protocol feature pauses the PCA infusion. This capability deactivates the patient dose button and continuous infusion when the hospital-set respiratory and/or oxygen-saturation rate alarm limits are exceeded.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Special

Physical Requirements

Width: 4.50 in (114 mm)	Left: N/A
Depth: 7.50 in (191 mm)	Right: N/A
Height: 15.00 in (381 mm)	Front: N/A
Max Weight: 6 lbs (2.7 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Powered via connection to Alaris PC Unit.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	NM
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	NM
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	NM
					Total:	6

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6891-005

Description: Pump, Infusion, Pump Module Set, Allowance
Manufacturer: BD - Alaris Infusion
Vendor: Unspecified
Model: TBD

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	Allowance placeholder - Pump Module not in AttaniaNM
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	Allowance placeholder - Pump Module not in AttaniaNM
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	Allowance placeholder - Pump Module not in AttaniaNM
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	Allowance placeholder - Pump Module not in AttaniaNM
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	Allowance placeholder - Pump Module not in AttaniaNM

Total: 10

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6891-005

Description: Pump, Infusion, Pump Module, Allowance
Manufacturer: BD - Alaris Infusion
Vendor: Unspecified
Model: TBD

Mfr #: TBD
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: N/A	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	3	Allowance placeholder - Pump Module not in AttaniaNM
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	Allowance placeholder - Pump Module not in AttaniaNM
Total:					4	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4182-008



Description: Pump, Infusion, Syringe
Manufacturer: BD - Alaris Infusion
Vendor: Unspecified
Model: Alaris Syringe Module (8110)

Mfr #: 8110
Vendor #:

Syringe infusion module. Attaches to Alaris PC Unit. Features Guardrails Suite MX software, flowrate from 0.01mL/hr to 999mL/hr, automatic syringe detection (syringes sized 1cc to 60cc) and set-based pressure sensing disc features.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 4.50 in (114 mm)	Left: N/A
Depth: 7.50 in (191 mm)	Right: N/A
Height: 15.00 in (381 mm)	Front: N/A
Max Weight: 5 lbs (2.3 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Powered via connection to Alaris PC Unit.
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1 NM	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1 NM	
Total:					2	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 3374-007



Description: Pump, Suction/Aspirator, General, Portable

Manufacturer: Armstrong Medical Industries ((800) 323-4220)

Vendor: SSCOR, Inc. ((818) 504-4054)

Model: SSCOR DUET (AE-6975)

Mfr #: AE-6975

Vendor #: AE-6975

Portable Suction/Aspirator. AC power w/ battery back-up, self contained charger, disposable canister, battery level indicator. 50-550mmHg performance. 50lpm airflow.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 115	Watts: 41
Hz: 60	Amps: 0.4
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 17.00 in (432 mm)	Left: N/A
Depth: 5.25 in (133 mm)	Right: N/A
Height: 9.00 in (229 mm)	Front: N/A
Max Weight: 11 lbs (5.0 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	227	Alcove, Code Cart	Project	Draft (New)	3	on crash cart

Total: 3

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4193-001



Description: Rack, Crutch/Cane/Walker

Manufacturer: Performance Health ((630) 393-6000)

Vendor: Unspecified

Mfr #: 7297

Vendor #:

Model: Walker Rack 7297

Wall mounted rack for walkers, crutches and/or canes. Holds up to 3 dozen pairs of crutches or six to eight walkers. Holds up to twelve 6" or 8" diameter rolls or six of the larger sizes. Includes 2 mounting brackets and four anchor bolts. Extends 16.5" from wall.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 1-Fixed	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/C	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 66.00 in (1676 mm)	Left: N/A
Depth: 16.50 in (419 mm)	Right: N/A
Height: 5.25 in (133 mm)	Front: N/A
Max Weight: 22 lbs (10.0 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	2	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	
					Total:	3

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 4248-000

Description: Regulator, Suction, Intermittent/Continuous
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 3.00 in (76 mm)	Left: N/A
Depth: 4.50 in (114 mm)	Right: N/A
Height: 5.50 in (140 mm)	Front: N/A
Max Weight: 1 lbs (0.5 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / Yes
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	on headwall
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	on headwall
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	on headwall
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	on headwall
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

Total: 41

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4266-051



Description: Scale, Clinical, Adult, Digital, Floor

Manufacturer: seca Corp ((410) 694-6980)

Vendor: Unspecified

Model: seca 769 Digital Column Scale w/BMI

Mfr #: 769 1321 994

Vendor #:

Digital column scale with height rod included. Capacity up to 450 lbs. and weight graduation of 0.2 lbs. Features display panel with brightly illuminated numbers, and casters for transport. Battery operation. Slip resistant cast iron base and anti-tip column design. Functions include TARE, HOLD, BMI, lbs/kg switch-over, and automatic / manual switch-off. Batteries are included. Power adapter optional.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 11.75 in (298 mm)	Left: N/A
Depth: 16.50 in (419 mm)	Right: N/A
Height: 32.75 in (832 mm)	Front: N/A
Max Weight: 14 lbs (6.4 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:
Battery powered
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	4H01	SP Corridor - Exam	Project	Draft (New)	1	
Standardized Patients	4H02	SP Corridor - Psy	Project	Draft (New)	1	
					Total:	2

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4277-011



Description: Scale, Clinical, Infant, Mobile

Manufacturer: Welch Allyn, Inc., A Hill-Rom Company ((800) 535-6663)

Vendor: Unspecified

Mfr #: 4802D-AX-XB /48800-1

Vendor #:

Model: Scale-Tronix 4802D Standard w/ 1 gm Weigh & Cart

Pediatric/infant scale on stainless steel cart. Configured with 1 gram weighing option. For daily and serial weighing of neonates, infants, toddlers and diapers. Weighing range up to 45 lb./ 20kg. Heavy-duty, stainless steel, portable unit with a built-in digital readout. Features automatic zero, weight recall, reweigh. Battery powered (6 "D" size alkaline batteries). Batteries and Power Adapter included. Removable acrylic cradle: 25" x 14". Cart is mounted on large rubber casters and comes with tissue dispenser. Made in USA.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 36
Hz: 60	Amps: 0.3
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 33.00 in (838 mm)	Left: N/A
Depth: 20.00 in (508 mm)	Right: N/A
Height: 33.00 in (838 mm)	Front: N/A
Max Weight: 60 lbs (27.2 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5427-035



Description: Scale, Clinical, Infant/Pediatric, Digital

Manufacturer: seca Corp ((410) 694-6980)

Vendor: Unspecified

Model: seca 334 Mobile Digital

Mfr #: 334 1321 008

Vendor #:

Digital infant scale. Features 44 lbs. capacity. Lightweight (6.2 lbs.) and handle for transportation. Functions include automatic switch-off, TARE, auto-hold, BMIF, and kg/lbs. switch-over and automatic switch-off. Graduation: 0.5 oz / 5 g < 10 kg > 10 g. Battery operated. A/C UL adapter is optional.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 25.25 in (641 mm)	Left: N/A
Depth: 12.00 in (305 mm)	Right: N/A
Height: 4.25 in (108 mm)	Front: N/A
Max Weight: 7 lbs (3.2 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:
Power supply: Batteries. Power adapter optional.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Existing (Reuse)	Draft (New)	1	Sits on instrument table
Total:					1	

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4941-005



Description: Shelving, Allowance, Supply, High Density (Movable)

Manufacturer: InnerSpace Corporation ((800) 467-7224)

Mfr #: QWFLTR (Custom)

Vendor: InnerSpace Corporation ((800) 467-7224)

Vendor #: QWFLTR (Custom)

Model: QuickWire QWFLTR

Storage with Floor Track System. For general/ bulk storage items in any facility. Compact wire shelving units on a low-profile floor track system (minimum 3 ft. aisle required). Easily move multiple QuickWire shelving units when units are placed back-to-back and/or side-to-side on one rolling carriage. Features 4in.H x 2in.W V groove track, aluminum extrusion with hardened steel construction, custom built manual carriages capable of carrying loads of up to 2500 lbs, aluminum rails, reinforced nylon wheel and articulating mid-wheel assemblies & molded ledges. Carriages consist of four rails and four corner assemblies on which 1-4 shelving units can be placed. Includes anchoring hardware Molded ledges& end stop kits.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/V	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: REM	Left: N/A
Depth: REM	Right: N/A
Height: REM	Front: N/A
Max Weight: REM	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:
Final size and weight vary. System is customized per customer requirements.

- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	Allowance estimate for custom equipment

Total: 1

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4298-010



Description: Shelving, Wire, Chrome, 48 inch

Manufacturer: InnerSpace Corporation ((800) 467-7224)

Vendor: InnerSpace Corporation ((800) 467-7224)

Model: Quickwire QWP5247

Mfr #: QWP5247

Vendor #: QWP5247

Chrome wire shelving unit with 5 shelves. 48in.W x 24in.D x 74in.H. Features (4) 1 inch diameter chrome posts, (4) 1/4 inch chrome wire shelves, plastic shelf sleeves. Maximum load capacity 1200 lbs (545.5 kg). Accessories such as shelf ledges, dividers, plastic shelf mats and label holders available.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 48.00 in (1219 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 74.00 in (1880 mm)	Front: N/A
Max Weight: 96 lbs (43.5 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: Yes	Pre-approval:
---------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	4	
Storage	437	S&T Workroom/Storage	Project	Draft (New)	4	
Storage	467	SIM Equipment Storage	Project	Draft (New)	4	

Total: 12

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4360-042



Description: Stand, IV, Stainless Steel
Manufacturer: Blickman Inc. ((800) 247-5070)
Vendor: Unspecified
Model: 1415SS-4 (5-Leg, 4-Hook)

Mfr #: 0531415400
Vendor #: 0531415400

Stainless steel IV stand with 4 hooks and 24" diameter, 5-leg polished die cast aluminum base with Wall Saver end caps. 2-inch hard rubber swivel casters. Tru-Loc friction knob height adjustment 52" - 94".

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 24.00 in (610 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 94.00 in (2388 mm)	Front: N/A
Max Weight: 11 lbs (5.0 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	8	1 IV stands designated for the Alaris Pump Controller
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	2	1 IV stands designated for the Alaris Pump Controller
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	2	1 IV stands designated for the Alaris Pump Controller
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	36	2 IV stands designated for the Alaris Pump Controller
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	7	2 IV stands designated for the Alaris Pump Controller

Total: 55

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4414-052



Description: Stool, Exam, Cushion-Seat

Manufacturer: Midmark Corporation ((800) 643-6275)

Vendor: Midmark Corporation ((800) 643-6275)

Model: Ritter 272 Value Series

Mfr #: 272-001

Vendor #: 272-001

Cushion seat exam stool with pneumatic height adjustment. Lever hand control adjusts seat height 18 to 24.5 inches. Five-leg, black composite caster base. Weight capacity 225 lbs.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 23.00 in (584 mm)	Left: N/A
Depth: 23.00 in (584 mm)	Right: N/A
Height: 24.50 in (622 mm)	Front: N/A
Max Weight: 15 lbs (6.8 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	6	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	1	
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	

Total: 21

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4416-040



Description: Stool, Step

Manufacturer: Clinton Industries, Inc. ((800) 441-9131)

Vendor: Unspecified

Model: T-40 Chrome

Mfr #: T-40

Vendor #:

Chrome step stool. All welded steel construction, rubber tread top, and reinforced rubber feet. 350 lb. load capacity.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 14.25 in (362 mm)	Left: N/A
Depth: 11.25 in (286 mm)	Right: N/A
Height: 9.00 in (229 mm)	Front: N/A
Max Weight: 9 lbs (4.1 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	4	

Total: 4

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4433-003



Description: Stretcher, Crib, Pediatric
Manufacturer: Pedigo Products, Inc ((800) 822-3501)
Vendor: Unspecified
Model: 500 Pediatric Crib/Stretcher

Mfr #: 500-1
Vendor #:

Transport crib/stretcher with two-inch bar spacing and epoxy finish. Features one-handed rail adjustment, total accessibility, two-caster locking system, manual Fowler backrest, Trendelenburg/reverse Trend. Foam pad mattress with staph-chek cover. Optional security top. Made in USA.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 30.25 in (768 mm)	Left: N/A
Depth: 60.00 in (1524 mm)	Right: N/A
Height: 59.50 in (1511 mm)	Front: N/A
Max Weight: 254 lbs (115.2 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Height is 76-3/4" with security top in place
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	specified by Jane Kleinman

Total: 1

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 5082-017



Description: Stretcher, Procedure, EYE/ ENT, Chair

Manufacturer: TransMotion Medical, Inc. (866-860-8447)

Vendor: Unspecified

Model: TMM5-FB Surgical Stretcher-Chair

Mfr #: TMM5-FB

Vendor #:

Surgical Stretcher-Chair; Features 100% motorized positioning from chair to full stretcher; 500 lb. patient weight capacity; 24 inch wide patient surface; 25-33 inch Seat Height Range; 24VDC full battery powered with 110V/220V charger; Two 24VDC batteries; Articulating Headpiece; Quick Release Backrest for Emergency CPR Function; Chrome plated, padded, side rails with Tuck-Away feature; 4 inch Memory Foam Cushions; 4-Way Stretch Blue Fusion Fabric - Standard; Folding Footrest. Additional options & accessories available. Call for volume discounts.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: 120
Hz: 60	Amps: 1
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 30.25 in (768 mm)	Left: N/A
Depth: 72.50 in (1842 mm)	Right: N/A
Height: 33.00 in (838 mm)	Front: N/A
Max Weight: 250 lbs (113.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	

Total: 4

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4429-058



Description: Stretcher, Transport

Manufacturer: MAC Medical, Inc. ((877) 828-9975)

Vendor: Unspecified

Model: PTF1000 (5th Wheel Transport Stretcher)

Mfr #: PTF1000

Vendor #:

5th wheel transport stretcher. Features retractable center fifth wheel for effortless directional movement and cornering, dual Fowler assisted life back rest with 90° adjustability, Trendelenburg & reverse Trendelenburg +/- 18°, dual pedestal lift mechanism, 4" comfort mattress, 2 hook removable telescopic IV pole, O2 storage, 8" Tente total locking casters, 6 brake/steer pedals, 750 lb. weight capacity. Height adjustable range: 22"-35" from bottom of mattress. Overall dimensions: 83 5/16" long x 33 9/16" wide. Patient surface: 75 7/16" long x 29 1/8" wide.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 33.75 in (857 mm)	Left: N/A
Depth: 83.50 in (2121 mm)	Right: N/A
Height: 51.00 in (1295 mm)	Front: N/A
Max Weight: 400 lbs (181.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Height includes side rails.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Existing (Reuse)	Draft (New)	2	
Storage	467	SIM Equipment Storage	Project	Draft (New)	2	
Total:					4	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 9600-019



Description: Table, Exam/Treatment, Manual Adjust, Electric

Manufacturer: Midmark Corporation ((800) 643-6275)

Vendor: Midmark Corporation ((800) 643-6275)

Model: Ritter 204 (Dwr Htr, Pelvic Tilt, Seamless Top)

Mfr #: 204-002/ 002-0871

Vendor #: 204-002/ 002-0871

Exam table with seamless upholstery top, drawer heater, pelvic tilt and air spring assist manual back adjustment. Features stirrups, electrical receptacle, storage drawers with pass-thru, paper roll holder. 500 lb. weight capacity.

General Product Detail

Arch Sig: Yes	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 115	Watts: 633
Hz: 60	Amps: 5.5
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 30.00 in (762 mm)	Left: N/A
Depth: 76.00 in (1930 mm)	Right: N/A
Height: 33.25 in (845 mm)	Front: N/A
Max Weight: 220 lbs (99.8 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Existing (Reuse)	Draft (New)	100	Qty: 10 Existing

Total: 100

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5912-051



Description: Table, Instrument, 30-36 inch
Manufacturer: Lakeside Manufacturing, Inc. ((800) 558-8565)
Vendor: Unspecified
Model: 8357 (36in.W x 24in.D)

Mfr #: 8357
Vendor #: 8357

Stainless steel instrument table with under shelf. Legs and frame are rugged 19 gauge square tubing. Shelves are turned down 1 inch and hemmed back 3/4 inch. 36in.W x 24in.D x 34in.H.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 36.00 in (914 mm)	Left: N/A
Depth: 24.00 in (610 mm)	Right: N/A
Height: 34.00 in (864 mm)	Front: N/A
Max Weight: 56 lbs (25.4 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470A	Patient Flex, Maternity / OB	Existing (Reuse)	Draft (New)	1	Infant scale sits on top
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	

Total: 2

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 5934-014



Description: Table, Overbed, General

Manufacturer: Stryker Medical ((269) 329-2100)

Vendor: Unspecified

Model: Tru-Fit Single Top w/o Vanity

Mfr #: 3150-000-100

Vendor #: 3150-000-100

Single-top overbed table with 17.5in x 31in top surface. No vanity. Low profile, U-shaped base, height adjustable. Safe working load 25 lbs. Specify laminate, edge/base colors when ordering.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 32.00 in (813 mm)	Left: N/A
Depth: 21.25 in (540 mm)	Right: N/A
Height: 43.75 in (1111 mm)	Front: N/A
Max Weight: 43 lbs (19.5 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	30	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

Total: 41

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 7607-006



Description: Table, Therapy, Massage, Powered

Manufacturer: EarthLite ((760) 599-1112)

Vendor: Unspecified

Model: Ellora Vista Lift (30"W)

Mfr #: ELLORA-30

Vendor #:

30" wide, single pedestal, powered massage table (without headrest). Height range 18"-37", foot control. Pro-Push Deluxe 3 layer 3" cushioning, rounded salon corners. Lift capacity 650 lbs. (Bariatric rated). Optional tilt, salon top, headrest. ADA compliant. Made in the USA with components sourced globally. UL/CE listed.

General Product Detail

Arch Sig: No	Spatially Sig: Yes
Arch Code: 2-Movable, Elect	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 110	Watts: 330
Hz: 60	Amps: 3
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 30.00 in (762 mm)	Left: N/A
Depth: 73.00 in (1854 mm)	Right: N/A
Height: 37.00 in (940 mm)	Front: N/A
Max Weight: 180 lbs (81.6 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

- Specification:**
UL/CE listed.
Adjustable height range 18"-37".
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	437	S&T Workroom/Storage	Existing (Reuse)	Draft (New)	1	
Total:					1	

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 6097-000

Description: Telephone, Desktop
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/V	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: 115	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: N/A	Left: N/A
Depth: N/A	Right: N/A
Height: N/A	Front: N/A
Max Weight: N/A	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Administration	484	Office, Director - Administration	Project	Draft (New)	0	
Administration	482	Office, General - Administration	Project	Draft (New)	0	
Administration	480	Office, Tech	Project	Draft (New)	0	
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	0	

Total: 0

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6531-000

Description: Telephone, Desktop, Digital, Single Line
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/V	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 7.00 in (178 mm)	Left: N/A
Depth: 6.00 in (152 mm)	Right: N/A
Height: 6.50 in (165 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470K	Control Room	Project	Draft (New)	0	5 desk phones (1 phone per station)

Total: 0

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6529-000

Description: Telephone, Wall
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/V	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 4.50 in (114 mm)	Left: N/A
Depth: 4.50 in (114 mm)	Right: N/A
Height: 8.50 in (216 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470	Patient Flex Corridor	Project	Draft (New)	0	Wall phone in PF area at each entry
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	0	@ nurse station
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	0	@ nurse station
Storage	437	S&T Workroom/Storage	Project	Draft (New)	0	

Total: 0

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 8149-000

Description: Telephone, Wall, Digital, Single Line
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 6-IT/Computers	ADA: Yes
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/V	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 7.00 in (178 mm)	Left: N/A
Depth: 6.25 in (159 mm)	Right: N/A
Height: 6.50 in (165 mm)	Front: N/A
Max Weight: 3 lbs (1.4 kg)	Back: N/A
Mounting: Wall	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: Yes	

Product and Project Item Notes

Structural Requirements

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:**
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470D, 470,G, 470H, 475	Debrief Room	Project	Draft (New)	0	
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	0	

Total: 0

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet



Description: Ultrasound, Imaging, Handheld
Manufacturer: GE Healthcare - Imaging Systems (800-886-0815)
Vendor: Unspecified
Model: Vscan

Atta ID: 4919-014

Mfr #: H45571CC

Vendor #:

Handheld Ultrasound. Provides black and white anatomic and color-coded blood flow images in real-time. Device is complemented with Vscan Gateway software enabling image management at customers PC, a Vscan web portal including online access to product and clinical information. Vscan is cleared for cardiac, abdominal, urology, fetal /obstetrics, pediatric, selected peripheral vascular applications, as well as for thoracic / pleural motion and fluid detection. Following components are included in standard configuration: Vscan scanner with probe, docking station including global AC adapter with interchangeable plugs (USA and China, EU, UK, Australia) and USB cable, Vscan Gateway software, one rechargeable battery, 4 GB Micro-SD memory card, soft case, and access to Vscan web portal with entitlement to full online education package. In some countries gel (60 g bottle) may also be included in the standard configuration.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 3.00 in (76 mm)	Left: N/A
Depth: 5.50 in (140 mm)	Right: N/A
Height: 1.25 in (32 mm)	Front: N/A
Max Weight: 2 lbs (0.9 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Dimensions are for handheld scanner only. System also includes docking station for battery charging.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	1 NM	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4589-048



Description: Ventilator, Adult / Pediatric
Manufacturer: Vyaire Medical
Vendor: Unspecified
Model: VELA Plus

Mfr #: 16531-00
Vendor #:

Adult/pediatric ventilator with A/C, SIMV, CPAP/PSV/NPPV, PRVC modes, active & passive humidification. Volume, pressure breath types and apnea backup. Inspiratory flow support, exhalation valve provides accurate volume delivery. Parallel printer port, VGA output, nurse call port and data output port.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 13.00 in (330 mm)	Left: N/A
Depth: 14.50 in (368 mm)	Right: N/A
Height: 12.00 in (305 mm)	Front: N/A
Max Weight: 38 lbs (17.2 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Requires 40-85 psig for high pressure and 0.5 psig max (0-80 L/min) for low pressure.
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	4	*Order with rolling stand accessory, per Vuong request

Total: 4

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4589-048



Description: Ventilator, Adult / Pediatric
Manufacturer: Vyaire Medical
Vendor: Vyaire Medical
Model: VELA Plus

Mfr #: 16531-00
Vendor #: 16531-00

Adult/pediatric ventilator with A/C, SIMV, CPAP/PSV/NPPV, PRVC modes, active & passive humidification. Volume, pressure breath types and apnea backup. Inspiratory flow support, exhalation valve provides accurate volume delivery. Parallel printer port, VGA output, nurse call port and data output port.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 120	Watts: N/A
Hz: 60	Amps: N/A
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 13.00 in (330 mm)	Left: N/A
Depth: 14.50 in (368 mm)	Right: N/A
Height: 12.00 in (305 mm)	Front: N/A
Max Weight: 38 lbs (17.2 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

- Specification:**
Requires 40-85 psig for high pressure and 0.5 psig max (0-80 L/min) for low pressure.
- Structural:**
- Electrical:**
- Plumbing:**
- Mechanical:**

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	2	*Order with rolling stand accessory, per Vuong request
Total:					2	

FKP CannonDesign
UTHSC SON Simulation Center
Specification Coversheet

Atta ID: 6740-000

Description: Walker, Bariatric
Manufacturer: Unspecified
Vendor: Unspecified
Model: Unspecified

Mfr #:
Vendor #:

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 22.50 in (572 mm)	Left: N/A
Depth: 15.50 in (394 mm)	Right: N/A
Height: 41.00 in (1041 mm)	Front: N/A
Max Weight: 9 lbs (4.1 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Seismic: N	Pre-approval:
-------------------	----------------------

- Specification:
- Structural:
- Electrical:
- Plumbing:
- Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	4	QTY: 4 - 1 per ST areaNM
Storage	467	SIM Equipment Storage	Project	Draft (New)	2	NM 2 per PF Room

Total: 6

UTHSC SON Simulation Center

Specification Coversheet

Atta ID: 4661-010



Description: Warmer, Infant, Care System

Manufacturer: GE Healthcare - Maternal/Infant Care ((800) 345-2700)

Mfr #: 6650-0004-901

Vendor: Unspecified

Vendor #:

Model: Giraffe OmniBed (E&W Port/N&S Grommet)

Infant Care System. In bed scale w/ weight trending, tubing grommet on south and north wall. Rail mounting system, Baby Susan rotating mattress, dual access doors, East and West portholes, pass-thru door w/ sliding organizer, x-ray cassette tray, radiant warmer, 19" mattress. Elevating base (58"-70", canopy down), servo controlled humidifier, bed tilting 0-12 degrees Trendelenberg/Reverse Trendelenberg.

General Product Detail

Arch Sig: Yes	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 115	Watts: 1035
Hz: 60	Amps: 9
Phase: Single	BTU/hr: N/A
KVA:	Ded. Circuit: No
Emer. Power: Yes	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 45.00 in (1143 mm)	Left: N/A
Depth: 26.00 in (660 mm)	Right: N/A
Height: 94.00 in (2388 mm)	Front: N/A
Max Weight: 304 lbs (137.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: Yes
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Height reflects maximum with canopy up, maximum height with canopy down: 70".

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	

Total: 1

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4657-021



Description: Warmer, Patient, Hypothermia

Manufacturer: 3M Infection Prevention Division ((651) 733-1110)

Mfr #: 77500

Vendor: Unspecified

Vendor #:

Model: Bair Hugger 775

Forced-air patient warming unit. Two airflow settings, three hose-end temperature sensors. Tracks and reports forced-air warming therapy duration, over-temperature information, calibration, built-in hour meter, and fault codes via front panel.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 2-Movable, Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: 110	Watts: 1550
Hz: 60	Amps: 11.7
Phase: Single	BTU/hr: 1600
KVA:	Ded. Circuit: No
Emer. Power: No	Plug Type: Type B (NEMA 5-15)

Physical Requirements

Width: 14.00 in (356 mm)	Left: N/A
Depth: 13.00 in (330 mm)	Right: N/A
Height: 13.00 in (330 mm)	Front: N/A
Max Weight: 16 lbs (7.3 kg)	Back: N/A
Mounting: Counter/Cart/Table/Pole	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

BTU's = 1,600 high/1,330 low

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	
					Total:	1

FKP CannonDesign

UTHSC SON Simulation Center

Specification Coversheet



Description: Waste Can, 20-31 Gallon
Manufacturer: Rubbermaid Commercial Products (800-347-9800)
Vendor: Unspecified
Model: 3569-88 Untouchable Square (Black, 23 gal)

Atta ID: 5407-030

Mfr #: FG356988BLA

Vendor #:

Square waste container without lid. 23 gal capacity and integrated liner cinch. Black. Ships 4/case.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 15.50 in (394 mm)	Left: N/A
Depth: 16.50 in (419 mm)	Right: N/A
Height: 31.00 in (787 mm)	Front: N/A
Max Weight: 22 lbs (10.0 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Ships 4 per case.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	1	
Storage	437	S&T Workroom/Storage	Project	Draft (New)	1	
Storage	467	SIM Equipment Storage	Project	Draft (New)	1	

Total: 3

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4687-069



Description: Waste Can, Bio-Hazardous

Manufacturer: Grainger (800-472-4643)

Vendor: Unspecified

Model: Tough Guy 6GAJ6 (8 gal. Red)

Mfr #: 6GAJ6

Vendor #: 6GAJ6

Biohazardous waste can. 8 gallon capacity. Red. Features plastic construction, square shape and pedal-operated lid.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 16.25 in (413 mm)	Left: N/A
Depth: 16.25 in (413 mm)	Right: N/A
Height: 18.25 in (464 mm)	Front: N/A
Max Weight: 7 lbs (3.2 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	1	1 designated for the entire PF Room
Total:					1	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4688-028



Description: Waste Can, Open Top

Manufacturer: Rubbermaid Commercial Products (800-347-9800)

Vendor: Unspecified

Model: 2543 Fire Resistant Black (28 qt.)

Mfr #: FG254300BLA

Vendor #: FG254300BLA

28 qt., open top, fire resistant (fiberglass) wastebasket with rounded corners. CSFM and UL approved for fire safety. Black.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 0-Unassigned	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 10.50 in (267 mm)	Left: N/A
Depth: 14.50 in (368 mm)	Right: N/A
Height: 15.50 in (394 mm)	Front: N/A
Max Weight: 4 lbs (1.8 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Patient Care Flex	470K	Control Room	Project	Draft (New)	4	

Total: 4

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4688-028



Description: Waste Can, Open Top

Manufacturer: Rubbermaid Commercial Products (800-347-9800)

Mfr #: FG254300BLA

Vendor: Unspecified

Vendor #: FG254300BLA

Model: 2543 Fire Resistant Black (28 qt.)

28 qt., open top, fire resistant (fiberglass) wastebasket with rounded corners. CSFM and UL approved for fire safety. Black.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 10.50 in (267 mm)	Left: N/A
Depth: 14.50 in (368 mm)	Right: N/A
Height: 15.50 in (394 mm)	Front: N/A
Max Weight: 4 lbs (1.8 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Administration	484	Office, Director - Administration	Existing (Reuse)	Draft (New)	1	Existing FTE's
Administration	482	Office, General - Administration	Existing (Reuse)	Draft (New)	2	
Administration	482	Office, General - Administration	Project	Draft (New)	1	New FTE
Administration	480	Office, Tech	Project	Draft (New)	2	
Patient Care Flex	470D, 470,G, 470H, 475	Debrief Room	Project	Draft (New)	4	
Patient Care Flex	470E	Nursing Alcove	Project	Draft (New)	1	
Patient Care Flex	470B, 470C, 470I, 470J	Patient Flex, ICU	Project	Draft (New)	4	
Patient Care Flex	470A	Patient Flex, Maternity / OB	Project	Draft (New)	1	
Patient Care Flex	470L	Patient Flex, NICU	Project	Draft (New)	1	
Skills and Tasks	440,450, 4H07	Skills & Tasks, Bed (10)	Project	Draft (New)	36	
Skills and Tasks	460	Skills & Tasks, Bed (5)	Project	Draft (New)	5	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 4920-117



Description: Waste Can, Step-On

Manufacturer: Rubbermaid Commercial Products (800-347-9800)

Vendor: Unspecified

Model: 1883609 Slim Jim Resin Front Step 8 Gal./Black

Mfr #: 1883609

Vendor #: 1883609

Step-on waste can with 8 gallon (30L) capacity. Black. Features plastic construction, quiet and controlled lid closure, internal hinge design, liner retainer band.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Non-Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 16.75 in (425 mm)	Left: N/A
Depth: 10.75 in (273 mm)	Right: N/A
Height: 21.25 in (540 mm)	Front: N/A
Max Weight: 6 lbs (2.7 kg)	Back: N/A
Mounting: Floor	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Product and Project Item Notes

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Specification:

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Standardized Patients	401 - 421 (Odd numbers)	Exam Room	Project	Draft (New)	10	
Standardized Patients	423, 425, 427, 429	Psy Consultation	Project	Draft (New)	4	
Total:					14	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6292-017



Description: Wheelchair, Adult, Bariatric

Manufacturer: Drive DeVilbiss Healthcare ((866) 694-5085)

Vendor: Unspecified

Model: Sentra EC 22in. [STD22ECDFA-SF]

Mfr #: STD22ECDFA-SF

Vendor #:

Heavy-duty adult wheelchair. Features 22 inch seat width, detachable full arms, dual cross brace, swing away footrest, height adjustable seat 17.5-19.5 inches. Carbon steel frame, chrome hand rims, push to lock wheel locks, padded arm rests. Weight capacity 450 lbs.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 29.00 in (737 mm)	Left: N/A
Depth: 42.00 in (1067 mm)	Right: N/A
Height: 36.00 in (914 mm)	Front: N/A
Max Weight: 57 lbs (25.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Folded width is 13.5 inches.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	467	SIM Equipment Storage	Project	Draft (New)	2	2 per PF Room
Total:					2	

**UTHSC SON Simulation Center
Specification Coversheet**

Atta ID: 6292-017



Description: Wheelchair, Adult, Bariatric

Manufacturer: Drive DeVilbiss Healthcare ((866) 694-5085)

Vendor: Drive DeVilbiss Healthcare ((866) 694-5085)

Model: Sentra EC 22in. [STD22ECDFA-SF]

Mfr #: STD22ECDFA-SF

Vendor #: STD22ECDFA-SF

Heavy-duty adult wheelchair. Features 22 inch seat width, detachable full arms, dual cross brace, swing away footrest, height adjustable seat 17.5-19.5 inches. Carbon steel frame, chrome hand rims, push to lock wheel locks, padded arm rests. Weight capacity 450 lbs.

General Product Detail

Arch Sig: No	Spatially Sig: No
Arch Code: 3-Movable, Non-Elect	ADA: No
Custom Code: Unassigned	Antimicrobial: No
Furnish Install: O/O	Type: Medical
	Green: No

Electrical Requirements

Volts: N/A	Watts: N/A
Hz: N/A	Amps: N/A
Phase: N/A	BTU/hr: N/A
KVA: N/A	Ded. Circuit: No
Emer. Power: No	Plug Type: N/A

Physical Requirements

Width: 29.00 in (737 mm)	Left: N/A
Depth: 42.00 in (1067 mm)	Right: N/A
Height: 36.00 in (914 mm)	Front: N/A
Max Weight: 57 lbs (25.9 kg)	Back: N/A
Mounting: Floor-Mobile	Top: N/A
	Bottom: N/A

Utility and Technology Requirements

Water - Cold: No	Gasses: No
Water - Hot: No	Drain: No
Water - Treated: No	Steam: No
Vent: No	Vacuum - Dental / Medical: No / No
Tech Connect: No	

Structural Requirements

Seismic: No	Pre-approval:
--------------------	----------------------

Product and Project Item Notes

Specification:

Folded width is 13.5 inches.

Structural:

Electrical:

Plumbing:

Mechanical:

Location

Department	Room#	Room	Funding Source	Item Status	Qty	Item Notes
Storage	435	Bulk Storage, HD Shelving	Project	Draft (New)	4	

Total: 4